

31

31

AOV INTERNATIONAL LLP
MAINTENANCE OF BIOMEDICAL EQUIPMENT
SERVICE REPORT



HOSPITAL NAME	CHC Sidhwan bet	COMPLAINT NUMBER	HMF
HOSPITAL TYPE	CHC	COMPLAINT RECEIVED DATE / TIME	
DISTRICT	Ludhiana	COMPLAINT ATTENDED DATE / TIME	11-7-22
USER CONTACT DETAILS		COMPLAINT RECTIFIED DATE / TIME	11-7-22
EQUIPMENT TYPE (TICK THE BOX)	<input type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input checked="" type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	BARCODE NUMBER :	4120110041	MANUFACTURER :	MEDICAL SYSTEMS
	EQUIPMENT NAME :	ECG MACHINE	MODEL NO :	CARDIOFAX 3R
	DEPARTMENT :	EMERGENCY	SERIAL NUMBER :	F10451009

PROBLEM IDENTIFIED	<p>* Machine Heating up badly. * Leads Not Working. * Printer Not Working. * Bulb electrodes Damaged. * Roller Not Working.</p>
--------------------	---


ACTION TAKEN	<p>* Machine checked properly and found that the machine is heating badly and even not working. Bulb electrodes and patient cable are not working and found damaged. Machine is not in repairable condition. Hence, recommend for condemned.</p>
--------------	--

PHASE - NEUTRAL (VOLT):
 PHASE - EARTH (VOLT):
 NEUTRAL - EARTH (VOLT):

ENGINEER'S REMARKS	<p>Machine is Non-Repairable. So, Recommend for Condemnation.</p>
--------------------	---

SPARES CONSUMED OR REQUIRED	PART NAME	QTY	SERIAL NO. / LOT NO.

CUSTOMER REMARK IF ANY	H
------------------------	---

CUSTOMER SIGNATURE WITH SEAL	H	FOR AOV INTERNATIONAL LLP
		 ENGINEER NAME & SIGNATURE