

32

AOV INTERNATIONAL LLP MAINTENANCE OF BIOMEDICAL EQUIPMENT SERVICE REPORT



HOSPITAL NAME	sdh khanna	COMPLAINT NUMBER	HNF
HOSPITAL TYPE	sdh	COMPLAINT RECEIVED DATE / TIME	
DISTRICT	ludhiana	COMPLAINT ATTENDED DATE / TIME	5-8-2022
USER CONTACT DETAILS		COMPLAINT RECTIFIED DATE / TIME	5-8-2022
EQUIPMENT TYPE (TICK THE BOX)	<input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	BARCODE NUMBER :	3120320149	MANUFACTURER :	Wellgers.
	EQUIPMENT NAME :	ECG Machine	MODEL NO :	PISCES-A-103
	DEPARTMENT :	ECG Room	SERIAL NUMBER :	2K1207S21174-P

PROBLEM IDENTIFIED: Machine is not working

ACTION TAKEN: we checked & found. Machine Battery-faulty, Display faulty, keypad not working. some card faulty inside. Not a repairable condition. recommend to condemnation.

PHASE - NEUTRAL (VOLT): 236 ✓
 PHASE - EARTH (VOLT): 234 ✓
 NEUTRAL - EARTH (VOLT): 2V

ENGINEER'S REMARKS

SPARES CONSUMED OR REQUIRED	PART NAME	QTY	SERIAL NO. / LOT NO.

CUSTOMER REMARK IF ANY

CUSTOMER SIGNATURE WITH SEAL

 ਸ਼੍ਰੀਮੀਅਰ ਮੈਡੀਕਲ ਅਫਸਰ ਇੰ: ਸਿਵਲ ਹਸਪਤਾਲ, ਖੰਨਾ

ENGINEER NAME & SIGNATURE
 FOR AOV INTERNATIONAL LLP