

34

AOV INTERNATIONAL LLP

MAINTENANCE OF BIOMEDICAL EQUIPMENT

SERVICE REPORT



HOSPITAL NAME	Sdh Ichanna	COMPLAINT NUMBER	HNI
HOSPITAL TYPE	Sdh	COMPLAINT RECEIVED DATE / TIME	
DISTRICT	Ludhiana	COMPLAINT ATTENDED DATE / TIME	5-8-2022
USER CONTACT DETAILS		COMPLAINT RECTIFIED DATE / TIME	5-8-2022

EQUIPMENT TYPE (TICK THE BOX) UNDER CONTRACT UNDER WARRANTY UNDER AMC/CMC PHYSICAL DAMAGE CONDEMNED

EQUIPMENT DETAIL	MANUFACTURER :
BARCODE NUMBER : 3120320151	Schiller
EQUIPMENT NAME : ECG Machine	MODEL NO : Cardiovit-17-2
DEPARTMENT : ECG Room	SERIAL NUMBER : 020.01784

PROBLEM IDENTIFIED: Machine is not working properly.

ACTION TAKEN: And we checked & found machine keypad not repairable condition, machine meter board faulty, & machine body breakable. Not a machine not repairable condition. So, recommend to Condemnation.

■ PHASE - NEUTRAL (VOLT): 236V ■ PHASE - EARTH (VOLT): 234V ■ NEUTRAL - EARTH (VOLT): 232V

ENGINEER'S REMARKS

SPARES CONSUMED OR REQUIRED	PART NAME	QTY	SERIAL NO. / LOT NO.

CUSTOMER REMARK IF ANY

MSBhasin
 Bahinoleka
 ਸੀਨੀਅਰ ਮੈਡੀਕਲ ਔਫਸਰ ਇੰਚੋ:
 ਮੁਕਤ ਸਾਹਿਬ, ਖੰਨਾ
 CUSTOMER SIGNATURE WITH SEAL

Amul
 FOR AOV INTERNATIONAL LLP
 ENGINEER NAME & SIGNATURE