



AOV INTERNATIONAL LLP

MAINTENANCE OF BIOMEDICAL EQUIPMENT SERVICE REPORT



HOSPITAL NAME	Sdh Samrala	COMPLAINT NUMBER	HNF
HOSPITAL TYPE	Sdh	COMPLAINT RECEIVED DATE / TIME	
DISTRICT	Ludhiana	COMPLAINT ATTENDED DATE / TIME	14-7-2022
USER CONTACT DETAILS		COMPLAINT RECTIFIED DATE / TIME	14-7-2022
EQUIPMENT TYPE (TICK THE BOX)	<input type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	BARCODE NUMBER :	3120340051	MANUFACTURER :	Meditce England
	EQUIPMENT NAME :	Boyles apparatus	MODEL NO :	140129 ANESI-15M/36
	DEPARTMENT :	OT	SERIAL NUMBER :	110124

PROBLEM IDENTIFIED : Machine is not working.

ACTION TAKEN : Ser we check & found machine is not repairable condition. AS recommend to ~~condemned~~ Condemned.

PHASE - NEUTRAL (VOLT): 240 ✓
 PHASE - EARTH (VOLT): 236 ✓
 NEUTRAL - EARTH (VOLT): 4 V

ENGINEER'S REMARKS

SPARES CONSUMED OR REQUIRED	PART NAME	QTY	SERIAL NO. / LOT NO.
	/	/	/

CUSTOMER REMARK IF ANY

<p>CUSTOMER SIGNATURE WITH SEAL</p> <p style="text-align: center;">MEDICAL OFFICER Civil Hospital, Samrala 140722</p>	<p style="text-align: right;">FOR AOV INTERNATIONAL LLP</p> <p style="text-align: center;">ENGINEER NAME & SIGNATURE</p>
--------------------------------------------------------------------------------------------------------------------------------------	------------------------------------------------------------------------------------------------------------------------------