



36

36

SERVICE REPORT

HOSPITAL NAME	Dn Maluholee	COMPLAINT NUMBER	
HOSPITAL TYPE	DH	COMPLAINT RECEIVED DATE/TIME	13/12/22
DISTRICT	Maluholee	COMPLAINT ATTENDED DATE/TIME	13/12/22
USER CONTACT DETAILS		COMPLAINT RECTIFIED DATE/TIME	13/12/22
EQUIPMENT TYPE (TICK THE BOX)	<input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CNC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER: 2230040017	MANUFACTURER:
	EQUIPMENT NAME: OT (Hydraulic)	MODEL NO: NA
	DEPARTMENT: Surgery OT	SERIAL NUMBER: NA

PROBLEM IDENTIFIED
Table is not working

ACTION TAKEN
Check of fault that this table is not working from longtime updown position not working and very old model and fully rusted, not recommend to use in OT. So this table is not repairable.

PHASE - NEUTRAL (VOLT):
 PHASE - EARTH (VOLT):
 NEUTRAL - EARTH (VOLT):

ENGINEER'S REMARKS
Recommended for condemnation.

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.

CUSTOMER REMARKS IF ANY

JOB COMPLETED SATISFACTORY YES NO

Senior Medical Officer
Hospital Maluholee

FOR AOV INTERNATIONAL LLP

ENGINEER NAME & SIGNATURE
Jagjit Singh 9289690053