



SERVICE REPORT

HOSPITAL NAME	DH Malerkotla	COMPLAINT NUMBER	
HOSPITAL TYPE		COMPLAINT RECEIVED DATE/TIME	13/12/22
DISTRICT		COMPLAINT ATTENDED DATE/TIME	13/12/22
USER CONTACT DETAILS		COMPLAINT RECTIFIED DATE/TIME	13/12/22
EQUIPMENT TYPE (TICK THE BOX)	<input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CNC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER: 2230040013	MANUFACTURER:
	EQUIPMENT NAME: OT (Hydraulic)	MODEL NO: MA
	DEPARTMENT: Surgery OT	SERIAL NUMBER: MA

PROBLEM IDENTIFIED

ACTION TAKEN

OT table is not working
 Check and found that this table up down pump
 is not working, very old model nobody can repair
 this table, not recommended to use in OT. So this
 table is not repairable and recommended for condemnation.

PHASE - NEUTRAL (VOLT):
 PHASE - EARTH (VOLT):
 NEUTRAL - EARTH (VOLT):

ENGINEER'S REMARKS

Recommended for condemnation.

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.

CUSTOMER REMARKS IF ANY

JOB COMPLETED SATISFACTORY YES NO

CUSTOMER SIGNATURE WITH SEAL

Jas
 Senior Medical Officer
 DH Hospital Malerkotla

FOR AOV INTERNATIONAL LLP

Yogesh
 TAMS

ENGINEER NAME & SIGNATURE

AOV