

30

39

AOV INTERNATIONAL LLP

BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF PUNJAB HEALTH SYSTEMS CORPORATION



SERVICE REPORT

3296

| | | | |
|-------------------------------|---|--------------------------------|---------|
| HOSPITAL NAME | PHC Boothgarh | COMPLAINT NUMBER | MNF |
| HOSPITAL TYPE | PHC | COMPLAINT RECEIVED DATE/ TIME | |
| DISTRICT | Boothgarh (Mohali) | COMPLAINT ATTENDED DATE/ TIME | 21/9/22 |
| USER CONTACT DETAILS | 98722 64570 | COMPLAINT RECTIFIED DATE/ TIME | |
| EQUIPMENT TYPE (TICK THE BOX) | <input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CNC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED | | |

| | | | | |
|------------------|-------------------|--------------|----------------|-------------|
| EQUIPMENT DETAIL | EQUIPMENT NUMBER: | 51600 900 24 | MANUFACTURER: | Focus |
| | EQUIPMENT NAME: | Microscope | MODEL NO: | BIO supreme |
| | DEPARTMENT: | Lab | SERIAL NUMBER: | 23763 |
| | | | | |

PROBLEM IDENTIFIED
- Main head problem -
slide recent main all lens faulty

ACTION TAKEN
Check and found mic is very old model main head lens faulty and slide recent main all lens faulty. MIC is not repairable

PHASE - NEUTRAL (VOLT): PHASE - EARTH (VOLT): NEUTRAL - EARTH (VOLT):

ENGINEER'S REMARKS
Not Repairable

| PARTS USED | PART NAME | QTY. | SERIAL NO. / LOT NO. |
|------------|-----------|------|----------------------|
| | | | |
| | | | |
| | | | |

OTHER REMARKS

COMPLETED SATISFACTORY
 YES
 NO

CLIENT SIGNATURE WITH SEAL
(Signature)
 22/09/22
 ਮੁਖਿ ਮਹਿ (ਮੋਹਲੀ)

FOR AOV INTERNATIONAL LLP
(Signature)
Baljit
 ENGINEER NAME & SIGNATURE

