

40

# AOV INTERNATIONAL LLP

## BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF PUNJAB HEALTH SYSTEMS CORPORATION



### SERVICE REPORT

3297

HOSPITAL NAME	PHC Boothgarh	COMPLAINT NUMBER	HWF
HOSPITAL TYPE	PHC	COMPLAINT RECEIVED DATE/TIME	
DISTRICT	Boothgarh (Mohali)	COMPLAINT ATTENDED DATE/TIME	21/9/22
USER CONTACT DETAILS	9879244570	COMPLAINT RECTIFIED DATE/TIME	

EQUIPMENT TYPE (TICK THE BOX)  UNDER CONTRACT  UNDER WARRANTY  UNDER AMC/CNC  PHYSICAL DAMAGE  CONDEMNED

EQUIPMENT DETAIL	EQUIPMENT NUMBER: 5160090025	MANUFACTURER: FOCUS
	EQUIPMENT NAME: MICROSCOPE	MODEL NO: MPM2
	DEPARTMENT: Lab	SERIAL NUMBER: 000

PROBLEM IDENTIFIED	Slide resentment faulty transformer faulty. and main head lens
ACTION TAKEN	Check and found MIC is very old and main head lens and slide adjustment faulty MIC is not repairable

PHASE - NEUTRAL (VOLT):  PHASE - EARTH (VOLT):  NEUTRAL - EARTH (VOLT):

ENGINEER'S REMARKS	MIC is not repairable
--------------------	-----------------------

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.

CUSTOMER REMARKS IF ANY	
-------------------------	--

COMPLETED SATISFACTORY <input type="checkbox"/> YES <input type="checkbox"/> NO CUSTOMER SIGNATURE WITH SEAL <i>Asrar</i> 21/9/22 21/9/22	FOR AOV INTERNATIONAL LLP  ENGINEER NAME & SIGNATURE
---	--

