

41

41

# AOV INTERNATIONAL LLP

## BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF PUNJAB HEALTH SYSTEMS CORPORATION



### SERVICE REPORT

3298

HOSPITAL NAME	PHC Boothgarh.	COMPLAINT NUMBER	HNF
HOSPITAL TYPE	PHC	COMPLAINT RECEIVED DATE/TIME	
DISTRICT	Boothgarh (Mohali)	COMPLAINT ATTENDED DATE/TIME	21/9/22
USER CONTACT DETAILS	9872244570	COMPLAINT RECTIFIED DATE/TIME	

EQUIPMENT TYPE (TICK THE BOX)  UNDER CONTRACT  UNDER WARRANTY  UNDER AMC/CNC  PHYSICAL DAMAGE  CONDEMNED

EQUIPMENT DETAIL	EQUIPMENT NUMBER:	S160090026	MANUFACTURER:	Lebomed
	EQUIPMENT NAME:	MICROSCOPE	MODEL NO:	VISION 250
	DEPARTMENT:	Lab	SERIAL NUMBER:	4187

PROBLEM IDENTIFIED	Main head lens faulty slider adjustment faulty all lens
ACTION TAKEN	Check and found main head lens and slider adjustment faulty and has become faulty due to very old not repairable

PHASE - NEUTRAL (VOLT):  PHASE - EARTH (VOLT):  NEUTRAL - EARTH (VOLT):

ENGINEER'S REMARKS: MIC is not repairable.

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.

CUSTOMER REMARKS IF ANY

JOB COMPLETED SATISFACTORY  YES  NO

CUSTOMER SIGNATURE WITH SEAL: *[Signature]*  
 ਮਹਿੰਦਰ ਸਿੰਘ (ਮੋਹਾਲੀ)

FOR AOV INTERNATIONAL LLP: *[Signature]* *[Signature]*  
 ENGINEER NAME & SIGNATURE

