

48

48

AOV INTERNATIONAL LLP
MAINTENANCE OF BIOMEDICAL EQUIPMENT
SERVICE REPORT



HOSPITAL NAME	Civil Hospital Doda	COMPLAINT NUMBER	HNF
HOSPITAL TYPE	CHC Doda	COMPLAINT RECEIVED DATE / TIME	HNF
DISTRICT	Muktsar	COMPLAINT ATTENDED DATE / TIME	26.7.22
USER CONTACT DETAILS	81468-11699	COMPLAINT RECTIFIED DATE / TIME	26.7.22
EQUIPMENT TYPE (TICK THE BOX)	<input type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	BARCODE NUMBER :	4170020011	MANUFACTURER :	Reason Diagnostic
	EQUIPMENT NAME :	Photoelectric colorimeter	MODEL NO :	OCTA1
	DEPARTMENT :	Laboratory	SERIAL NUMBER :	18325

PROBLEM IDENTIFIED	Machine is not getting on
--------------------	---------------------------


ACTION TAKEN	visit the place and check the MLC and found that MLC power supply is faulty and Due to the MLC is more than 15 years old it is not possible to arrange power supply so that MLC is not in repair condition.
--------------	---


■ PHASE - NEUTRAL (VOLT): 230 ■ PHASE - EARTH (VOLT): 232 ■ NEUTRAL - EARTH (VOLT): 2

ENGINEER'S REMARKS			
--------------------	--	--	--


SPARES CONSUMED OR REQUIRED	PART NAME	QTY	SERIAL NO. / LOT NO.

CUSTOMER REMARK IF ANY	
------------------------	--


 CUSTOMER SIGNATURE WITH SEAL


 26-7-22

FOR AOV INTERNATIONAL LLP


 ENGINEER NAME & SIGNATURE