

43

AOV INTERNATIONAL LLP
MAINTENANCE OF BIOMEDICAL EQUIPMENT
SERVICE REPORT



HOSPITAL NAME	Civil Hospital Doda.	COMPLAINT NUMBER	HNF
HOSPITAL TYPE	CHC Doda.	COMPLAINT RECEIVED DATE / TIME	HNF
DISTRICT	Muketsar	COMPLAINT ATTENDED DATE / TIME	26.7.22
USER CONTACT DETAILS	81468-11699	COMPLAINT RECTIFIED DATE / TIME	26.7.22
EQUIPMENT TYPE (TICK THE BOX)	<input type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	BARCODE NUMBER :	4170020012	MANUFACTURER :	Electricity
	EQUIPMENT NAME :	Photoelectric colorimeter	MODEL NO :	N/A
	DEPARTMENT :	Laboratory	SERIAL NUMBER :	434

PROBLEM IDENTIFIED: Machine NOT getting ON.

ACTION TAKEN: Visit the place and check the m/c and found there is problem in power supply and test tube holder. ~~Due~~ to m/c is order then 15 years. So, it is not possible to have spares to repair it so that m/c is not repairable.

■ PHASE - NEUTRAL (VOLT): 230 ■ PHASE - EARTH (VOLT): 232 ■ NEUTRAL - EARTH (VOLT): 2

ENGINEER'S REMARKS

SPARES CONSUMED OR REQUIRED	PART NAME	QTY	SERIAL NO. / LOT NO.

CUSTOMER REMARK IF ANY

CUSTOMER SIGNATURE WITH SEAL: 26-7-22
 ENGINEER NAME & SIGNATURE:

FOR AOV INTERNATIONAL LLP