



AOV INTERNATIONAL LLP
BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF
PUNJAB HEALTH SYSTEMS CORPORATION 1383



SERVICE REPORT

HOSPITAL NAME	District hospital	COMPLAINT NUMBER	HNF
HOSPITAL TYPE		COMPLAINT RECEIVED DATE/ TIME	HNF
DISTRICT	PH	COMPLAINT ATTENDED DATE/ TIME	29-8-22
USER CONTACT DETAILS	nikhar	COMPLAINT RECTIFIED DATE/ TIME	

EQUIPMENT TYPE (TICK THE BOX) UNDER CONTRACT UNDER WARRANTY UNDER AMC/CNC PHYSICAL DAMAGE CONDEMNED

EQUIPMENT DETAIL	EQUIPMENT NUMBER: 2170040086	MANUFACTURER: Cognate
	EQUIPMENT NAME: Infusion apparatus ^{Infusion pump}	MODEL NO: N/A
	DEPARTMENT: Gynae OT	SERIAL NUMBER: N/A

PROBLEM IDENTIFIED: MIC not working

ACTION TAKEN: Check the MIC and found older than 15 years so it is not in repairable condition.

PHASE - NEUTRAL (VOLT): PHASE - EARTH (VOLT): NEUTRAL - EARTH (VOLT):

ENGINEER'S REMARKS

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.

CUSTOMER REMARKS IF ANY

JOB COMPLETED SATISFACTORY <input type="checkbox"/> YES <input type="checkbox"/> NO	FOR AOV INTERNATIONAL LLP
CUSTOMER SIGNATURE WITH SEAL <i>Masrullah</i>	ENGINEER NAME & SIGNATURE <i>[Signature]</i>

