

46

46

# AOV INTERNATIONAL LLP

## BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMPP) OF PUNJAB HEALTH SYSTEMS CORPORATION

1357



### SERVICE REPORT

COMPLAINT NO.	31702	COMPLAINT NUMBER	1111
COMPLAINT TYPE	CDM	COMPLAINT RECEIVED DATE/TIME	11-11-22
CITY	Mudke	COMPLAINT ATTENDED DATE/TIME	21-11-22
COMPLAINT DETAILS		COMPLAINT RECEIVED DATE/TIME	24-11-22

UNDER CONTRACT   
  UNDER WARRANTY   
  UNDER MAINT.   
  PHYSICAL DAMAGE   
  CONSUMED

EQUIPMENT NO.	EQUIPMENT NUMBER	31702	MANUFACTURER	Toshiba
	EQUIPMENT NAME	Envi Auto Analyser	MODEL NO.	Chem 54
	DEPARTMENT	Laboratory	SERIAL NUMBER	10302

**PROBLEM IDENTIFIED**  
M/c Not working.

**ACTION TAKEN**  
Check the m/c and found that older than 15 years  
So its difficult to arrange spare of the m/c  
So that m/c is not in repairable condition.


PHASE-NEUTRAL (VOLT): 230   
  PHASE-EARTH (VOLT): 238   
  NEUTRAL-EARTH (VOLT): 0


**ENGINEER'S REMARKS**  
Nothing suggested

SPARE CONSUMED	PART NAME	QTY	SERIAL NO / LOT NO.

**CUSTOMER REMARKS**  
if any

WORK COMPLETED SATISFACTORILY   
  YES   
  NO

FOR AOV INTERNATIONAL LLP  
  
 ENGINEER NAME & SIGNATURE

  
 AOV