

47

47

AOV INTERNATIONAL LLP

MAINTENANCE OF BIOMEDICAL EQUIPMENT

SERVICE REPORT



HOSPITAL NAME	CHC SAROYA	COMPLAINT NUMBER	HNF
HOSPITAL TYPE	CHC	COMPLAINT RECEIVED DATE / TIME	17/8/22
DISTRICT	NALWANSAR	COMPLAINT ATTENDED DATE / TIME	17/8/22
USER CONTACT DETAILS		COMPLAINT RECTIFIED DATE / TIME	
EQUIPMENT TYPE (TICK THE BOX)	<input type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	BARCODE NUMBER : 4180050028	MANUFACTURER :
	EQUIPMENT NAME : BOYLES APPRATUS	MODEL NO :
	DEPARTMENT : MAIN OT	SERIAL NUMBER :

PROBLEM IDENTIFIED: BOYLES APPRATUS Recommended for Condemnation

ACTION TAKEN: Checked the MACHINE and found that it's condition is very poor Body has been rusted and very old MACHINE. SO it's non-repairable and Recommended for condemnation

PHASE - NEUTRAL (VOLT):
 PHASE - EARTH (VOLT):
 NEUTRAL - EARTH (VOLT):

ENGINEER'S REMARKS

SPARES CONSUMED OR REQUIRED	PART NAME	QTY	SERIAL NO. / LOT NO.

CUSTOMER REMARK IF ANY

CUSTOMER SIGNATURE WITH SEAL

Mahal Medical Officer
 Taphra Community Health Centre
 SAROYA (Nalwanshar)

Harpreet Singh
 ENGINEER NAME & SIGNATURE

FOR AOV INTERNATIONAL LLP