

48

48

# AOV INTERNATIONAL LLP

## MAINTENANCE OF BIOMEDICAL EQUIPMENT

### SERVICE REPORT



HOSPITAL NAME	CHC CHAROTA	COMPLAINT NUMBER	HNF
HOSPITAL TYPE	CHC	COMPLAINT RECEIVED DATE / TIME	09/08/2024
DISTRICT	BATHANKOT	COMPLAINT ATTENDED DATE / TIME	09/08/2024 14:51
USER CONTACT DETAILS		COMPLAINT RECTIFIED DATE / TIME	09/08/2024 00:00

EQUIPMENT TYPE (TICK THE BOX)  UNDER CONTRACT  UNDER WARRANTY  UNDER AMC/CMC  PHYSICAL DAMAGE

EQUIPMENT DETAIL	BARCODE NUMBER : 419020011	MANUFACTURER : other
	EQUIPMENT NAME : Examination mobile light	MODEL NO : NA
	DEPARTMENT : Main OT	SERIAL NUMBER : NA

PROBLEM IDENTIFIED: Mobile light not working.

ACTION TAKEN: Checked the MIC and found the age of equipment is more than 40 years and not repairable. So recommended for condemnation.

PHASE - NEUTRAL (VOLT):  PHASE - EARTH (VOLT):  NEUTRAL - EARTH (VOLT):

ENGINEER'S REMARKS

SPARES CONSUMED OR REQUIRED	PART NAME	QTY	SERIAL NO. / LOT NO

CUSTOMER REMARK IF ANY

CUSTOMER SIGNATURE WITH SEAL:

ENGINEER NAME & SIGNATURE:

FOR AOV INTERNATIONAL