

50

# AOV INTERNATIONAL LLP MAINTENANCE OF BIOMEDICAL EQUIPMENT SERVICE REPORT

HMF



HOSPITAL NAME	CMC Badshapur	COMPLAINT NUMBER	
HOSPITAL TYPE	CMC	COMPLAINT RECEIVED DATE / TIME	
DISTRICT	Patiala	COMPLAINT ATTENDED DATE / TIME	13/7/22
CONTACT DETAILS	9877740775	COMPLAINT RECTIFIED DATE / TIME	13/7/22
EQUIPMENT TYPE (TICK THE BOX)	<input type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	BARCODE NUMBER :	4140020010	MANUFACTURER :	Other
	EQUIPMENT NAME :	Microscope	MODEL NO :	DM-9
	DEPARTMENT :	Laboratory	SERIAL NUMBER :	N/A

PROBLEM IDENTIFIED

ACTION TAKEN

Very obsolete model and approximate 20 years old. Not in used more 10 years. The machine all lenses faulty. Machine in very poor condition, so recommended to condemnation.

■ PHASE - NEUTRAL (VOLT): 235    ■ PHASE - EARTH (VOLT): 231    ■ NEUTRAL - EARTH (VOLT): 1.5

ENGINEER'S REMARKS

This microscope is non-repairable. So, recommend to condemnation.

SPARES CONSUMED OR REQUIRED

PART NAME	QTY	SERIAL NO. / LOT NO.
/	/	/

CUSTOMER REMARK IF ANY

Satisfied with work. Operation behaviour.

FOR AOV INTERNATIONAL LLP

CUSTOMER SIGNATURE WITH SEAL  
*Serjeet*  
13/7/2022  
Senior Medical Officer  
C.H.C. Badshapur  
Mansindhan (R)

ENGINEER NAME & SIGNATURE  
*Rimki*