

51

51

AOV INTERNATIONAL LLP

MAINTENANCE OF BIOMEDICAL EQUIPMENT HNF

SERVICE REPORT



HOSPITAL NAME	CHC Badshapur	COMPLAINT NUMBER	
HOSPITAL TYPE	CHC	COMPLAINT RECEIVED DATE / TIME	8/7/22
DISTRICT	Patiala	COMPLAINT ATTENDED DATE / TIME	9/7/22
BRIEF CONTACT DETAILS	7347581278	COMPLAINT RECTIFIED DATE / TIME	
EQUIPMENT TYPE (TICK THE BOX)	<input type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	MANUFACTURER :
BARCODE NUMBER :	Elpro GE
EQUIPMENT NAME :	MODEL NO :
X-Ray Machine 60 MA	ST-60
DEPARTMENT :	SERIAL NUMBER :
X-Ray	C/56877

PROBLEM IDENTIFIED

ACTION TAKEN

This Machine has 30 years old model and obsolete. There are no spare parts available. This Machine cannot pass QA terms and conditions, so this Machine is recommended for condemnation.

PHASE - NEUTRAL (VOLT): 235
 PHASE - EARTH (VOLT): 233
 NEUTRAL - EARTH (VOLT): 1.5

ENGINEER'S REMARKS

This Machine is non-repairable. So, recommended for condemnation.

SPARES CONSUMED OR REQUIRED	PART NAME	QTY	SERIAL NO. / LOT NO.
	/		

CUSTOMER REMARK IF ANY

Cooperative behavior satisfied with work.

CUSTOMER SIGNATURE WITH SEAL Senior Medical Officer CHC, Badshapur	FOR AOV INTERNATIONAL LLP ENGINEER NAME & SIGNATURE
--	--