

SERVICE REPORT

3650

52

COMPLAINT NAME	CHC Dushan Sachan	COMPLAINT NUMBER	011115
CONTRACT TYPE	CHC	COMPLAINT RECEIVED DATE/ TIME	
DISTRICT	Patiala	COMPLAINT ATTENDED DATE/ TIME	14-07-2022
CONTRACT DETAILS		COMPLAINT RECTIFIED DATE/ TIME	
EQUIPMENT TYPE (TICK THE BOX)	<input type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CNC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER	4140040014	MANUFACTURER	CHC rest
	EQUIPMENT NAME	Centrifuge	MODEL NO	N4
	DEPARTMENT	Lab	SERIAL NUMBER	N4

PROBLEM IDENTIFIED

ACTION TAKEN

check and found that is very obsolete model no spare available in the market and very old model so this centrifuge is recommended for condemnation




PHASE - NEUTRAL (VOLT)
 PHASE - EARTH (VOLT)
 NEUTRAL - EARTH (VOLT)

ENGINEER'S REMARKS

Recommended for condemnation

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.

CUSTOMER REMARKS IF ANY

JOB COMPLETED SATISFACTORY <input type="checkbox"/> YES <input type="checkbox"/> NO CUSTOMER SIGNATURE WITH SEAL 	FOR AOV INTERNATIONAL LLP  ENGINEER NAME & SIGNATURE 
---	---