



# AOV INTERNATIONAL LLP

## BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF PUNJAB HEALTH SYSTEMS CORPORATION SERVICE REPORT



23870

HOSPITAL NAME	DH Kapurthala	COMPLAINT NUMBER	PB11395
HOSPITAL TYPE	DH	COMPLAINT RECEIVED DATE/TIME	18-9-23
DISTRICT	Kapurthala	COMPLAINT ATTENDED DATE/TIME	20-9-23
USER CONTACT DETAILS	9872211234	COMPLAINT RECTIFIED DATE/TIME	16-10-23
EQUIPMENT TYPE (TICK THE BOX)	<input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CNC <input type="checkbox"/> PHYSICAL DAMAGE <input checked="" type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER: 2110060035	MANUFACTURER: ESA OTC
	EQUIPMENT NAME: Ultrasound Machine	MODEL NO: MY LAB 40
	DEPARTMENT: Ultrasound Room	SERIAL NUMBER: 201481


**PROBLEM IDENTIFIED**  
Keypad Misfunctioning. Not Working Properly.



**ACTION TAKEN**  
As per OEM This Model is obsolete & out of production so company will not be able to provide the spare part of machine so we suggest this machine to be condemned.

PHASE - NEUTRAL (VOLT):     
 PHASE - EARTH (VOLT):     
 NEUTRAL - EARTH (VOLT):

**ENGINEER'S REMARKS**  
Machine suggested for condemnation

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.
	N/A		

**CUSTOMER REMARKS IF ANY**  
  
Senior Medical Officer  
Incharge Civil Hospital,  
Kapurthala

<b>JOB COMPLETED SATISFACTORY</b> <input type="checkbox"/> YES <input type="checkbox"/> NO	<b>FOR AOV INTERNATIONAL LLP</b>  Sugnu Kumar ENGINEER NAME & SIGNATURE
<b>CUSTOMER SIGNATURE WITH SEAL</b>	



Esaote Asia Pacific Diagnostic Private Limited

Corporate Office:  
Galaxy Business Park A-44 & 45  
Tower - C, Ground Floor,  
Sector - 62, Noida - 201 301  
Uttar Pradesh, India  
Tel : +91 120 471 2444, 475 9140 47  
Fax : +91 120 475 0148  
Toll Free : 1800 - 309 - 0234  
Web: www.esaote.com


Registered Office:  
17B-171, First Floor, Prajapat Nagar,  
Gautam Nagar, New Delhi - India - 110049  
CIN No: UR5105012007PTC170446

### END OF LIFE OF ESAOTE USG MACHINES

Dear sir,

This is to inform you that the below machine are out of production being obsolete model so company will not be able to provide the spare parts for the same

1. Esaote My Lab 40 with sr. no. 40-12712
2. Esaote My Lab 40 with sr. no. 30646044
3. Esaote My Lab 40 with sr. no. 4012455



Kindest regards

Sanjay Kumar

Manager-Technical Service (India and Subcontinent)

Esaote Asia Pacific Diagnostic Pvt. Ltd

