

AOV INTERNATIONAL LLP

**BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF
PUNJAB HEALTH SYSTEMS CORPORATION**



SERVICE REPORT

30356



HOSPITAL NAME	UCMC Basti Gujjan	COMPLAINT NUMBER	PB/2891
HOSPITAL TYPE	UCMC	COMPLAINT RECEIVED DATE/ TIME	31/10/23
DISTRICT	Jalandhar	COMPLAINT ATTENDED DATE/ TIME	31/10/23
USER CONTACT DETAILS		COMPLAINT RECTIFIED DATE/ TIME	31/10/23
EQUIPMENT TYPE (TICK THE BOX)	<input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER : 8100200001	MANUFACTURER : L&T
	EQUIPMENT NAME : Ultrasound Machine	MODEL NO : Sonalisa 32
	DEPARTMENT : Ultrasound	SERIAL NUMBER : 406605179

PROBLEM IDENTIFIED : Ultrasound Probe Faulty

ACTION TAKEN : Checked & found Ultrasound probe faulty & ~~checked & found~~ As per OGM This model is End of Life & service support Not available so we suggest this machine for condemnation.

PHASE - NEUTRAL (VOLT) :
 PHASE - EARTH (VOLT) :
 NEUTRAL - EARTH (VOLT) :

ENGINEER'S REMARKS : Machine suggested for Condemnation

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.
	N/A		/

CUSTOMER REMARKS IF ANY :

JOB COMPLETED SATISFACTORY YES NO

CUSTOMER SIGNATURE WITH SEAL : *Dev*
 Medical Officer
 U.C.M.C Basti Guzan,
 Jalandhar

FOR AOV INTERNATIONAL LLP
Sushil
SUSHIL KUMAR
 ENGINEER NAME & SIGNATURE



Ref:MS/CHD/31082023/1

Dt:31/08/2023

AOV International LLP
#986,First Floor, Sec-82
Mohali-160055.

Sub:End of life Certificate /Condemnation Certificate

Dear Sir/Madam,

It is to inform you that following Medical equipment supplied / installed at your institute By M/s Larsen & Toubro Limited is obsolete & service support not available, hence it is Fit for Condemnation.

S. No.	Equipment / Model Name / Serial No.
1	Ultrasound machine / Sonalisa 32 / Y06W5179

Thanking you,
For Skanray Technologies Ltd.



Mayank Srivastav
Sr. Service Engineer
Mohali(Punjab)