



# AOV INTERNATIONAL LLP

## BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF PUNJAB HEALTH SYSTEMS CORPORATION



30327

HOSPITAL NAME	SDH NAKODAR	COMPLAINT NUMBER	PB-12959
HOSPITAL TYPE	SDH	COMPLAINT RECEIVED DATE/TIME	02/11/23
DISTRICT	Jalandhar	COMPLAINT ATTENDED DATE/TIME	02/11/23
USER CONTACT DETAILS	9855478816	COMPLAINT RECTIFIED DATE/TIME	02/11/23
EQUIPMENT TYPE (TICK THE BOX)	<input type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input checked="" type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER:	3100170001	MANUFACTURER:	BPL
	EQUIPMENT NAME:	ECG machine	MODEL NO:	Cardiact 6208 view
	DEPARTMENT:	ECG ROOM	SERIAL NUMBER:	DETE7D7047

PROBLEM IDENTIFIED: machine not working.

ACTION TAKEN: Checked and found machine main Card or Acqui Card and display faulty. Asset repairing cost is more than ~~the~~ total Asset value. We suggest for this machine for RBER.

PHASE - NEUTRAL (VOLT):     
 PHASE - EARTH (VOLT):     
 NEUTRAL - EARTH (VOLT):

ENGINEER'S REMARKS: We suggest for RBER

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.
	/	/	/

CUSTOMER REMARKS IF ANY: N/A

JOB COMPLETED SATISFACTORY  YES  NO  
 CUSTOMER SIGNATURE WITH SEAL: **SURVINDER KAUR**  
 Sr. Pharmacy Officer  
 SDH, Nakodar

FOR AOV INTERNATIONAL LLP  
 ENGINEER NAME & SIGNATURE: **KAPIL KUMAR**

**Fwd: quotation for ecg machine 6208 view and 6108t**

1 message

Fri, Oct 27, 2023 at 9:39 AM

TRC TRC <trc.punjab@aovinternational.net>  
To: Sushil Kumar <sushil.zhjalandhar@aovinternational.net>

fyi



Thanks & Regards

**RAVI BHARDWAJ**

TRC ENGINEER

AOV International LLP

#986, First Floor, Sector - 82,

Mohali - 160055 (India)

Mob +91-9289690097,9816411427

Email trc.punjab@aovinternational.net

----- Forwarded message -----  
From: TRC TRC <trc.punjab@aovinternational.net>  
Date: Fri, Oct 6, 2023 at 11:36 AM  
Subject: Fwd: quotation for ecg machine 6208 view and 6108t  
To: SUNIL RANA <store.pb@aovinternational.net>, SHIV KUMAR <shivkumarnangal@gmail.com>  
Cc: Renu Dheer <dph.punjab@aovinternational.net>, Rahul Dadhwal <rahul.d@aovinternational.net>, Sumit Sumit <store.punjab@aovinternational.net>, Vikas Tomer <technicalhead.punjab@aovinternational.net>

Dear Sir,

Please find forwarded mail and Go Ahead for RBER .



Thanks & Regards

**RAVI BHARDWAJ**

TRC ENGINEER

AOV International LLP

#986, First Floor, Sector - 82,

Mohali - 160055 (India)

Mob +91-9289690097,9816411427

Email trc.punjab@aovinternational.net

----- Forwarded message -----  
From: SUNIL RANA <store.pb@aovinternational.net>  
Date: Wed, Sep 27, 2023 at 5:57 PM  
Subject: Re: quotation for ecg machine 6208 view and 6108t  
To: TRC TRC <trc.punjab@aovinternational.net>  
Cc: Rahul Dadhwal <rahul.d@aovinternational.net>, Devesh Kumar <devesh@aovinternational.net>

Dear TRC Team

Go Ahead FOR RBER

On Wed, Sep 27, 2023 at 5:45 PM Devesh Kumar <devesh@aovinternational.net> wrote

Dear Sunil Ji,

Please find details for RBER.

Asset Name: ECG MACHINE
Asset Make : BPL
Asset Model: 6108T
Asset Serial No: AVMR0J9246

Barcode: 5200150005

Quote Ref. No.	Description	Date	Qty	Price	Amount
Email: 27.09.2023	MAIN + MICON PCB ASSEMBLY 6108T	27-09-2023	1	12500	12500
Email: 27.09.2023	BATTERY 9.6V,1500MAH NIMH-DIGI/6108T	27-09-2023	1	2500	2500
Email: 27.09.2023	GEAR ASSEMBLY-108TDIGI	27-09-2023	1	940	940
<b>Total Spend Value</b>					<b>15940</b>
<b>Asset Value</b>					<b>25625</b>
<b>% of Spend Value</b>					<b>62%</b>

Asset Name: ECG MACHINE

Asset Make : BPL

Asset Model: CardIart 6208

Asset Serial No: DETE7D7047

Barcode: 3100170001

Quote Ref. No.	Description	Date	Qty	Price	Amount
Email: 27.09.2023	6208VIEW MAIN BOARD VER1.0 COLOUR DISPLAY	27-09-2023	1	20850	20850
Email: 27.09.2023	6208VIEW ECG PCB-NEW :1.0 VER	27-09-2023	1	16240	16240
<b>Total Spend Value</b>					<b>37090</b>
<b>Asset Value</b>					<b>25625</b>
<b>% of Spend Value</b>					<b>145%</b>

On Wed, Sep 27, 2023 at 3:02 PM SUNIL RANA <store.pb@aovinternational.net> wrote:  
Please Provide Barcode.

----- Forwarded message -----

From: Parmesh Kumar &lt;parmesh.kumar@bpl.in&gt;

Date: Wed, Sep 27, 2023 at 2:55 PM

Subject: quotation for ecg machine 6208 view and 6108t

To: Technical Support &lt;technicalhead.punjab@aovinternational.net&gt;

Cc: &lt;store.pb@aovinternational.net&gt;

Dear Sir,

We have checked the Both Ecg machine and both machine are defective

So please approve the below price of ecg machine spare required

Ecg 6108T below spare required (machine s.no --AVMK8J9246)

S.no	Material description	Model Number	Total price	Gst @	Total
1	MAIN + MICON PCB ASSEMBLY 6108T	ECG6108T	12,500	12%	14000
2	BATTERY 9.6V,1500MAH NIMH-DIGI/6108T	Ecg 6108 T	2500	28%	3200
3	GEAR ASSEMBLY-108TDIGI	ECG6108T	940	12%	1052

ECG 6208 View Ecg Machine Spare Required --- DETE7D7047

S.no	Material description	Model Number	Total price	Gst @	Total
1	6208VIEW MAIN BOARD VER1.0 COLOUR DISPLAY	ECG 6208 VIEW	20,850	18%	24850
2	6208VIEW ECG PCB-NEW :1.0 VER	6208VIEW	16240	18%	19163

Thanks &amp; Regards,

**Parmesh Kumar**


Lead Service Engineer


Ph: +91-9417869800 | Email: Parmesh.kumar@bpl.in

Ph: +91-80-2648 4388 | +91-80-2648 4350 | Toll Free 1800-425-2355


**BPL Medical Technologies Pvt. Ltd.**


11th KM, Bannerghatta Road, Arakere, Bangalore - 560 076, India

signature\_1463506506

signature\_3837076302

signature\_3749632969

signature\_2156186891

signature\_1516596612

signature\_1051074073

signature\_2526845643

• Cardiology • Critical Care & Surgery • Mother & Child Care • Imaging • Homecare • Consumables

Thanks & Regards

Sunil Rana  
AOV International LLP #986,  
First Floor, Sector - 82, Mohali - 160055  
(India) Mob : +91-9779885584  
Email: store.pb@aovinternational.net

Thanks & Regards

Sunil Rana  
AOV International LLP #986,  
First Floor, Sector - 82, Mohali - 160055  
(India) Mob : +91-9779885584  
Email: store.pb@aovinternational.net