

Ref:MS/CHD/09092023/1

Dt:09-09-2023

AOV International LLP  
#986,First Floor, Sec-82  
Mohali-160055.

**Sub:End of life Certificate /Condemnation Certificate**

Dear Sir/Madam,

It is to inform you that following Medical equipment supplied / installed at your institute  
By M/s Larsen & Toubro Limited is obsolete & service support not available, hence it is  
Fit for Condemnation.

S. No.	Equipment / Model Name / Serial No.
1	Cautery Machine / D400 / E0050173

Thanking you,  
For Skanray Technologies Ltd.



Mayank Srivastav  
Sr. Service Engineer  
Mohali( Punjab)





**BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF  
PUNJAB HEALTH SYSTEMS CORPORATION**



**AOV INTERNATIONAL LLP**

**SERVICE REPORT**

**22898**

HOSPITAL NAME		SDH Cuddesbaha		COMPLAINT NUMBER	PG12894
HOSPITAL TYPE		SDH		COMPLAINT RECEIVED DATE/TIME	31-10-25
DISTRICT		Ludhiana		COMPLAINT ATTENDED DATE/TIME	9-11-27
USER CONTRACT DETAILS		9419545297		COMPLAINT RECEIVED DATE/TIME	7-11-27
EQUIPMENT TYPE (TICK THE BOX)		<input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONCEALED			
EQUIPMENT DETAIL		EQUIPMENT NUMBER :	317026009	MANUFACTURER :	AGT Medical
		EQUIPMENT NAME :	Country machine	MODEL NO. :	V10 300S
		DEPARTMENT :	lab, OT	SERIAL NUMBER :	E0050129
PROBLEM IDENTIFIED	M/E not working.				
ACTION TAKEN	Checked the M/E and found M/E not getting ON so found that it main card faulty so card needs to replace but M/E is older than 25 years so spares not available.				
ENGINEER'S REMARKS	M/E recommended for condemnation, due to M/E has end of life				
SPARES CONSUMED	PART NAME	QTY.	SERIAL NO./LOT NO.		
	N/A				
CUSTOMER REMARKS IF ANY	M/E				
JOB COMPLETED SATISFACTORY		<input checked="" type="checkbox"/> YES <input type="checkbox"/> NO		FOR AOV INTERNATIONAL LLP	
CUSTOMER SIGNATURE WITH SEAL		 CUSTOMER SIGNATURE WITH SEAL		ENGINEER NAME & SIGNATURE  ENGINEER NAME & SIGNATURE	