

Bankers  
HDFCBank  
A/c. No. : 50200019052542  
IFSC CODE : HDFC0001391

APPROVED GOVT. CONTRACTOR  
GST NO : 03AAQPG3231K2ZU  
Drug Licence No. : PB-JA1-133380, PB-JA1-133381

Ph. (O) 2401950  
(O) 4605046, 47, 48  
Fax : 0181-4627240  
(Works) 0181-2912948  
Mobile : 98141-00448  
77104-00448

# MODERN Surgical House



SHOWROOM : G. T. Road, Near Distt. Jail, JALANDHAR-144 008  
Works : 1, Grovar Colony, 120 Ft Road, Opp. Adarsh Hospital, Jalandhar  
E-mail : shriyanmahajan804@gmail.com, officemodren@gmail.com Website : www.modernsurgicalhouse.com

Manufactures & Exporters of Surgical Instruments, Medical Equipment, Hospital Furniture, Lab Regents & Scientific Goods

Distributors for :

- |   |  |  |
|---|--|--|
| ❖ Johnson & Johnson Limited (Ethicon)   | ❖ J Mitra & Company Ltd. (Delhi)       | ❖ Qualikems Fine Chemicals Pvt. Ltd. (New Delhi) |
| ❖ L-P Int'l Ltd. (USA)                  | ❖ Johnson & Johnson Ltd. (New Delhi)   | ❖ Sutures India Pvt. Ltd. (Bangalore)            |
| ❖ Bayer Diagnostics India Ltd. (Baroda) | ❖ Cadila Hospital Products (Ahmedabad) | ❖ Birla 3M Ltd. (Bangalore)                      |
| ❖ Becton Dickinson India Ltd. (B.D.)    | ❖ Life Line System Pvt. Ltd. (Delhi)   | ❖ Hindustan Syringes & Medical Devices Ltd.      |
| ❖ Beacon Diagnostics                    | ❖ Vector Diagnostics                   | ❖ Span Diagnostics Ltd. (Ahmedabad)              |

Ref. No.....

Dated...14/12/23

Dear Sir,

THIS is to inform you that the Below Machine are out of production being obsolete model so company will not be able to provide the spare parts for the same.

1. Microcentrifuge All Micro (16 Tubes)

Kindest Regards

For Modern Surgical House

Dimpa

Prop.



# AOV INTERNATIONAL LLP



**BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF  
PUNJAB HEALTH SYSTEMS CORPORATION**



**37359**

## SERVICE REPORT

HOSPITAL NAME	SDH Nakodar	COMPLAINT NUMBER	PB14405
HOSPITAL TYPE	SDH	COMPLAINT RECEIVED DATE/ TIME	15/12/2023
DISTRICT	Jalandhar	COMPLAINT ATTENDED DATE/ TIME	15/12/23
USER CONTACT DETAILS		COMPLAINT RECTIFIED DATE/ TIME	15/12/2023
EQUIPMENT TYPE (TICK THE BOX)		<input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED	

EQUIPMENT DETAIL	EQUIPMENT NUMBER : 3100170012	MANUFACTURER : Shriyan Modern Surgical
	EQUIPMENT NAME : Centrifuge Machine	MODEL NO : ALMICRO (16 Tubes)
	DEPARTMENT : LAB	SERIAL NUMBER :

**PROBLEM IDENTIFIED** : Checked & found Armature & Motor Assembly faulty.

**ACTION TAKEN** : As per OEM This model is obsolete & The spare part is not available with company. So we suggest This machine for Condemnation

PHASE - NEUTRAL (VOLT) :     
  PHASE - EARTH (VOLT) :     
  NEUTRAL - EARTH (VOLT) :

**ENGINEER'S REMARKS** : Machine suggested for Condemnation @ LATTER Attached

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.
	N/A		

**CUSTOMER REMARKS IF ANY** :

JOB COMPLETED SATISFACTORY <input checked="" type="checkbox"/> YES <input type="checkbox"/> NO CUSTOMER SIGNATURE WITH SEAL 	FOR AOV INTERNATIONAL LLP Senior Medical Officer Hospital Nakodar SUSHEEL KUMAR ENGINEER NAME & SIGNATURE 
---	---

