



BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF
PUNJAB HEALTH SYSTEMS CORPORATION



INCIDENT REPORT

I. R. No. **111**
Date **12/12/23**
Region

1 Device Identification	
a. Asset No.	4200030038
b. Asset description	Oxygen Concentrator
c. Manufacturer Name	VOBEKAN
d. Model	Z7-5ZW
e. Serial No.	2105263221005
f. Any other Details	

2 Date & Location	
a. Hospital Name	CHC MORINDA
b. Department Name	Emergency
c. Date of Occurrence	12/12/23
d. Time of Occurrence	
e. Date of Reported	12/12/23
f. Time Reported	

3 Follow Up Action	
A. BM Work Order No.:	

3 Damage to Equipment	
A. Major <input checked="" type="checkbox"/>	B. Minor <input type="checkbox"/>

5 Description of the Incident (State exactly what was the sequence of events leading up to the incident and where the incident occurred Please add another paper if space provided is not enough)

Machine is submerged in flood & Rain water
Rusted badly & jammed. Not Repairable

6 Cause of Incident			
a. Physical Damage	<input type="checkbox"/>	j. Inadequate Illumination	<input type="checkbox"/>
b. Missing Spare/Parts	<input type="checkbox"/>	k. Hazardous Method or Procedure	<input type="checkbox"/>
c. Operating without Authorization	<input type="checkbox"/>	l. Non Use of Personal Protective Equipment	<input type="checkbox"/>
d. Wrong Usage / Setting	<input type="checkbox"/>	m. Patient Action	<input type="checkbox"/>
e. Unsafe Equipment	<input type="checkbox"/>	n. Fire, Explosion or Natural Calamities	<input checked="" type="checkbox"/>
f. Unsafe Position or Posture	<input type="checkbox"/>	o. Hazardous Personal Attire	<input type="checkbox"/>
g. Working on Moving or Dangerous Equipment	<input type="checkbox"/>	p. Unsafe Design or Arrangement	<input type="checkbox"/>
h. Distraction, Teasing or Willful Misconduct	<input type="checkbox"/>	q. Equipment Malfunction or Device Failure	<input type="checkbox"/>
i. Not Guarded or Improperly Guarded	<input type="checkbox"/>	Other Explain	<input type="checkbox"/>

7 Action Taken on Equipment Repair	
a) Minor problem and solved without any financial implications.	<input type="checkbox"/>
b) Minor problem and submitted the quote for required parts.	<input type="checkbox"/>
c) Major problem, needs vendor support for further trouble shooting.	<input type="checkbox"/>
d) Major problem. Can not be repaired, Proposed for BER	<input checked="" type="checkbox"/>

E) Replacement required parts details		Part Nos	Qty.	Approx. Cost
a)				
b)				
c)				
d)				
e)				
Total approx. estimate cost of Repair				

8 Remarks (if any)

Signatures	Engineer Name	<i>Ranjot Singh</i>	NODAL OFFICER	<i>[Signature]</i>
	Date		Date	ਸ਼੍ਰੀ ਮੈਡੀਕਲ ਅਫਸਰ ਏ. ਸੀ. ਐਚ. ਸੀ. ਮੋਰਿੰਡਾ, ਰੂਪਨਗਰ



AOV INTERNATIONAL LLP

**BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF
PUNJAB HEALTH SYSTEMS CORPORATION**



38401

SERVICE REPORT

HOSPITAL NAME	C.H.C MORINDA	COMPLAINT NUMBER	Service Call.
HOSPITAL TYPE	CHC	COMPLAINT RECEIVED DATE/ TIME	
DISTRICT	ROPAK	COMPLAINT ATTENDED DATE/ TIME	12/12/23
USER CONTACT DETAILS		COMPLAINT RECTIFIED DATE/ TIME	12/12/23

EQUIPMENT TYPE (TICK THE BOX)	<input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED			
EQUIPMENT DETAIL	EQUIPMENT NUMBER:	4200030038	MANUFACTURER:	YOBEKAN
	EQUIPMENT NAME:	Oxygen Concentrator.	MODEL NO:	ZV-52W
	DEPARTMENT:	Emergency.	SERIAL NUMBER:	2105263221005

PROBLEM IDENTIFIED: Due to flood & water machine is not working.

ACTION TAKEN: Checked & found the machine was submerged in rain & flood water. found moisture at base & wet. Main PCB, Compressor motor, dusted completely & jammed. All the filters are checked. All the repair parts are not repairable & also not available in market as m/c is china made.

PHASE - NEUTRAL (VOLT): 237
 PHASE - EARTH (VOLT): 235
 NEUTRAL - EARTH (VOLT): 02

ENGINEER'S REMARKS: This m/c is recommended for condemnation.

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.
	/	/	/

CUSTOMER REMARKS IF ANY:

JOB COMPLETED SATISFACTORY YES NO

FOR AOV INTERNATIONAL LLP
Ranbir Singh
Ravi Bhandwaj
 ENGINEER NAME & SIGNATURE

CUSTOMER SIGNATURE WITH SEAL
ਸ਼੍ਰੀ ਐਚ ਸੀ ਮੋਰਿੰਡਾ, ਰੁਪਨਗਰ

