



**BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF
PUNJAB HEALTH SYSTEMS CORPORATION**



INCIDENT REPORT

I. R. No. **122**
Date **12/12/23**
Region

1 Device Identification	
a. Asset No.	4200030079
b. Asset description	Oxygen Concentrator
c. Manufacturer Name	KONSUNG
d. Model	KSOC-10
e. Serial No.	Ma 21050581067
f. Any other Details	

2 Date & Location	
a. Hospital Name	CHC MORINDA
b. Department Name	General ward
c. Date of Occurrence	12/12/23
d. Time of Occurrence	
e. Date of Reported	12/12/23
f. Time Reported	

3 Follow Up Action	
A. BM Work Order No.:	

3 Damage to Equipment		
A. Major	<input checked="" type="checkbox"/>	B. Minor <input type="checkbox"/>

5 Description of the Incident (State exactly what was the sequence of events leading up to the incident and where the incident occurred Please add another paper if space provided is not enough)

*Machine is submerged in flood & Rain water
Rusted badly & jammed. Not Repairable.*

6 Cause of Incident			
a. Physical Damage	<input type="checkbox"/>	j. Inadequate Illumination	<input type="checkbox"/>
b. Missing Spare/Parts	<input type="checkbox"/>	k. Hazardous Method or Procedure	<input type="checkbox"/>
c. Operating without Authorization	<input type="checkbox"/>	l. Non Use of Personal Protective Equipment	<input type="checkbox"/>
d. Wrong Usage / Setting	<input type="checkbox"/>	m. Patient Action	<input type="checkbox"/>
e. Unsafe Equipment	<input type="checkbox"/>	n. Fire, Explosion or Natural Calamities	<input checked="" type="checkbox"/>
f. Unsafe Position or Posture	<input type="checkbox"/>	o. Hazardous Personal Attire	<input type="checkbox"/>
g. Working on Moving or Dangerous Equipment	<input type="checkbox"/>	p. Unsafe Design or Arrangement	<input type="checkbox"/>
h. Distraction, Teasing or Willfull Misconduct	<input type="checkbox"/>	q. Equipment Malfunction or Device Failure	<input type="checkbox"/>
i. Not Guarded or Temporarily Guarded	<input type="checkbox"/>	Other Explain	<input type="checkbox"/>

7 Action Taken on Equipment Repair	
a) Minor problem and solved without any financial implications.	<input type="checkbox"/>
b) Minor problem and submitted the quote for required parts.	<input type="checkbox"/>
c) Major problem, needs vendor support for further trouble shooting.	<input type="checkbox"/>
d) Major problem. Can not be repaired, Proposed for BER	<input checked="" type="checkbox"/>

E) Replacement required parts details		Part Nos	Qty.	Approx. Cost
a)				
b)				
c)				
d)				
e)				
Total approx. estimate cost of Repair				

8 Remarks (if any)

Signatures	Engineer Name	<i>Ranjit Singh</i>	NODAL OFFICER	<i>[Signature]</i>	
	Date		Date	ਸੀ. ਐਚ. ਸੀ. ਮਰਿੰਡਾ, ਰੁਪਨਗਰ	

AOV INTERNATIONAL LLP

BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF
PUNJAB HEALTH SYSTEMS CORPORATION



38407

SERVICE REPORT

HOSPITAL NAME	CHC MORINDA	COMPLAINT NUMBER	Service Call
HOSPITAL TYPE	CHC	COMPLAINT RECEIVED DATE/ TIME	
DISTRICT	ROPAR	COMPLAINT ATTENDED DATE/ TIME	12/12/23
USER CONTACT DETAILS		COMPLAINT RECTIFIED DATE/ TIME	12/12/23
EQUIPMENT TYPE (TICK THE BOX)	<input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER: 4200030079	MANUFACTURER: KONSUNG BIOMEDICAL
	EQUIPMENT NAME: OXYGEN CONCENTRATOR	MODEL NO: KOSC-10
	DEPARTMENT: GENERAL WARD	SERIAL NUMBER: Ma-2150581067

PROBLEM IDENTIFIED
Machine is not Working (flood & Rain Water submerged)

ACTION TAKEN
Checked and found the Machine was submerged in rain and flood water. In these machine the Compressor is jammed totally. and show the Error-3 on screen. and All the filter are checked. and repairable item are not Available in market Due to Chinese model.

PHASE - NEUTRAL (VOLT): 237
 PHASE - EARTH (VOLT): 235
 NEUTRAL - EARTH (VOLT): 02

ENGINEER'S REMARKS
This m/c is recommended for Condemnation.

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.
	/	/	/

CUSTOMER REMARKS IF ANY

JOB COMPLETED SATISFACTORY YES NO

FOR AOV INTERNATIONAL LLP
 Ravi Singh
 Kavi Bhardwa;
 ENGINEER NAME & SIGNATURE

CUSTOMER SIGNATURE WITH SEAL
 ਸੀ. ਐਚ. ਐਸ. ਮੋਰਿੰਡਾ
 ਡੀ. ਐਚ. ਐਸ. ਮੋਰਿੰਡਾ
 ਡੀ. ਐਚ. ਐਸ. ਮੋਰਿੰਡਾ

