

# AOV INTERNATIONAL LLP



**BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF  
PUNJAB HEALTH SYSTEMS CORPORATION**



**38626**

## SERVICE REPORT

HOSPITAL NAME	CIVIL HOSPITAL	COMPLAINT NUMBER	SERVICE CALL
HOSPITAL TYPE	DH	COMPLAINT RECEIVED DATE/ TIME	10/01/24
DISTRICT	PATHANKOT	COMPLAINT ATTENDED DATE/ TIME	10/01/24
USER CONTACT DETAILS	9815165400	COMPLAINT RECTIFIED DATE/ TIME	10/01/24.
EQUIPMENT TYPE (TICK THE BOX)	<input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER : 2190050399.	MANUFACTURER : MEDICO
	EQUIPMENT NAME : C-ARM machine.	MODEL NO : SUEGICO 60D
	DEPARTMENT : ORTHO OT	SERIAL NUMBER : 2304N.

**PROBLEM IDENTIFIED** : Machine Not working & Recommended for condemnation.

**ACTION TAKEN** : checked & found machine rusted & Not used from 7 years & all parts are damaged. So Recommended for condemnation.

PHASE - NEUTRAL (VOLT) :     
  PHASE - EARTH (VOLT) :     
  NEUTRAL - EARTH (VOLT) :

**ENGINEER'S REMARKS** : Recommended for condemnation to machine.

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.

**CUSTOMER REMARKS IF ANY** :

**JOB COMPLETED SATISFACTORY**     
  YES       NO     
 FOR AOV INTERNATIONAL LLP     
 RAI SINGH

**CUSTOMER SIGNATURE WITH SEAL** : *[Signature]*     
**ENGINEER NAME & SIGNATURE** : *[Signature]*