

AOV INTERNATIONAL LLP 5743

BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF  
PUNJAB HEALTH SYSTEMS CORPORATION



SERVICE REPORT

HOSPITAL NAME	DH Fatehgarh Sahib	COMPLAINT NUMBER	HINE
HOSPITAL TYPE	DH	COMPLAINT RECEIVED DATE/ TIME	
DISTRICT	Fatehgarh Sahib	COMPLAINT ATTENDED DATE/ TIME	
USER CONTACT DETAILS		COMPLAINT RECTIFIED DATE/ TIME	21/01/23
EQUIPMENT TYPE (TICK THE BOX)	<input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CNC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER: 2050670146	MANUFACTURER: Baxter
	EQUIPMENT NAME: Tube Sealer	MODEL NO: Hematron - III
	DEPARTMENT: Blood Bank	SERIAL NUMBER: 7423

PROBLEM IDENTIFIED: Tube Sealer is not properly working.

ACTION TAKEN: Checked & Found mother board PCB Card Faulty and tube Seal Hand piece Faulty. This spare part is not available in the market. So recommended for the Condemnation.

PHASE - NEUTRAL (VOLT):     
  PHASE - EARTH (VOLT):     
  NEUTRAL - EARTH (VOLT):

ENGINEER'S REMARKS

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.
	/	/	/

CUSTOMER REMARKS IF ANY

JOB COMPLETED SATISFACTORY <input type="checkbox"/> YES <input type="checkbox"/> NO CUSTOMER SIGNATURE WITH SEAL B.T.O., Civil Hospital, Fatehgarh Sahib	FOR AOV INTERNATIONAL LLP Ravinder Singh ENGINEER NAME & SIGNATURE
--	--

