

AOV INTERNATIONAL LLP

**BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF
PUNJAB HEALTH SYSTEMS CORPORATION**



31753

SERVICE REPORT

HOSPITAL NAME	Gurdaspur	COMPLAINT NUMBER	Offline
HOSPITAL TYPE	DH	COMPLAINT RECEIVED DATE/ TIME	
DISTRICT	Gurdaspur	COMPLAINT ATTENDED DATE/ TIME	16-3-24
USER CONTACT DETAILS	9413019853	COMPLAINT RECTIFIED DATE/ TIME	
EQUIPMENT TYPE (TICK THE BOX)	<input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input checked="" type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER: 2080120203	MANUFACTURER: Other.
	EQUIPMENT NAME: Operation table (H)	MODEL NO: NA
	DEPARTMENT: Main OT	SERIAL NUMBER: NA

PROBLEM IDENTIFIED: Machine not working.

ACTION TAKEN: Checked and found up-down movement faulty. This machine is in very poor condition and the matching part is not available in company. So we recommend this machine for condemnation.

PHASE - NEUTRAL (VOLT):
 PHASE - EARTH (VOLT):
 NEUTRAL - EARTH (VOLT):

ENGINEER'S REMARKS: Recommend for Condemnation

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.
	/		/

CUSTOMER REMARKS IF ANY:

JOB COMPLETED SATISFACTORY <input checked="" type="checkbox"/> YES <input type="checkbox"/> NO Senior Medical Officer Civil Hospital, Gurdaspur	FOR AOV INTERNATIONAL LLP Baljinder Singh ENGINEER NAME & SIGNATURE
CUSTOMER SIGNATURE WITH SEAL	