

AOV INTERNATIONAL LLP

**BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF
PUNJAB HEALTH SYSTEMS CORPORATION**



45216

HOSPITAL NAME	civil Hospital	COMPLAINT NUMBER	PB23309
HOSPITAL TYPE	SDH Dera Bassi	COMPLAINT RECEIVED DATE/ TIME	23/8/24
DISTRICT	Mohali	COMPLAINT ATTENDED DATE/ TIME	29/8/24
USER CONTACT DETAILS	9875962669	COMPLAINT RECTIFIED DATE/ TIME	
EQUIPMENT TYPE (TICK THE BOX)	<input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER: 3160160051	MANUFACTURER: Haier
	EQUIPMENT NAME: Deep Freezer -20	MODEL NO: HBD-286
	DEPARTMENT: PP unit	SERIAL NUMBER: B287N0068

PROBLEM IDENTIFIED: machine was not working.

ACTION TAKEN: check and found external body rusted and damaged, cooling circuit is clogged, compressor pressure is low and condenser fan is faulty. This is not repairable.

PHASE - NEUTRAL (VOLT): 227
 PHASE - EARTH (VOLT): 224
 NEUTRAL - EARTH (VOLT): 02

ENGINEER'S REMARKS: machine is recommended for condemnation

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.
	/	/	/

CUSTOMER REMARKS IF ANY: N/A

JOB COMPLETED SATISFACTORY <input type="checkbox"/> YES <input type="checkbox"/> NO CUSTOMER SIGNATURE WITH SEAL: <i>[Signature]</i>	FOR AOV INTERNATIONAL LLP ENGINEER NAME & SIGNATURE: <i>[Signature]</i>
---	--



4a



DEEP FREEZER

**DO NOT STORE
ANY VACCINE/DILUENT
IN DEEP FREEZER**

**DEEP FREEZER IS FOR
FREEZING ICE PACKS ONLY**



D.F.
3rd floor
BVS

