



Date: 01-10-2024

CHC Nurpur Bedi
Ropar Punjab


E-mail ID: store.pb@aovinternational.net

Dear Sir,

Sub:: Non Serviceability for ECG Machine (**Model-ECG-3Plus**) Equipment Serial Number- **E3P11100022** manufactured in 2011.

We hereby certify and declare that the aforesaid equipment has reached its end of life and no more repairs are applicable.

Thanks & Best Regards


Rajaram Krishnamurthy
National Service Manager



AOV INTERNATIONAL LLP



BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF
PUNJAB HEALTH SYSTEMS CORPORATION

45161



SERVICE REPORT

HOSPITAL NAME	CHC Nurgpur Bedi	COMPLAINT NUMBER	PB 23343
HOSPITAL TYPE	CHC	COMPLAINT RECEIVED DATE/ TIME	24/8/24
DISTRICT	ROPAR.	COMPLAINT ATTENDED DATE/ TIME	24/8/24
USER CONTACT DETAILS	9417585161	COMPLAINT RECTIFIED DATE/ TIME	5/10/24
EQUIPMENT TYPE (TICK THE BOX)	<input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER: 4200040004	MANUFACTURER: ASPEN
	EQUIPMENT NAME: ECG MIC.	MODEL NO: ECG-3
	DEPARTMENT: Emergency.	SERIAL NUMBER: E3P11100022

PROBLEM IDENTIFIED: Display not working.

ACTION TAKEN: Checked the machine found that Display faulty. The machine part is not available in company because this model are out of production and being obsolete, it's non repairable. So he Recommended for Condemnation.

PHASE - NEUTRAL (VOLT): 230
 PHASE - EARTH (VOLT): 225
 NEUTRAL - EARTH (VOLT): 5

ENGINEER'S REMARKS: Non Repairable.

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.
	/	/	/

CUSTOMER REMARKS IF ANY:

JOB COMPLETED SATISFACTORY <input type="checkbox"/> YES <input type="checkbox"/> NO CUSTOMER SIGNATURE WITH SEAL: <i>[Signature]</i> Senior Medical Officer CHC Nurgpur Bedi (Roopnagar)	FOR AOV INTERNATIONAL LLP <i>[Signature]</i> 9289690024 ENGINEER NAME & SIGNATURE
---	--

