



Ref: BMXI/EOL/STAD_PATIALA/GT-BLOT48/10/2024

END OF LIFE CERTIFICATE

Date – 10th Oct 2024

To,
STATE TRAINING AND DEMONSTRATION
Near Sherawala Gate, PATIALA, Punjab 147001

Dear Sir,

This is to certify that **GT-BLOT 48 (Serial Number: 1095-3)**; Qty. 1 No. has been installed in your hospital on **3/12/2013** has outlived its probable life and it is confirmed that the repair of the same is not cost effective as the machine is quite old. It is also informed to you that there is no buy-back cost for the same.

We, therefore, suggest you to for the condemnation of this equipment as per policy.

Thanking you and assuring you best services always.

Thanking You,
For bioMerieux India Pvt. Ltd.



**Bivash
CHAKRABORTY**

Digitally signed by Bivash
CHAKRABORTY
Date: 2024.10.10 15:08:57
+05'30'

Bivash CHAKRABORTY
Head- Regulatory Affairs & Quality
(Authorized Signatory)
Ph- 9810410025;
E-mail: bivash.chakraborty@biomerieux.com

bioMerieux India Pvt. Ltd.

43A – Okhla Industrial Estate – Phase III – Modi Mill Compound – New Delhi – 110020 – India
Tel.: +91(11) 42098800 – Fax: +91(11) 42098850 – www.biomerieuxindia.in
Toll Free No.: 18001027791 – customercare.india@biomerieux.com

AOV INTERNATIONAL LLP



**BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF
PUNJAB HEALTH SYSTEMS CORPORATION**



SERVICE REPORT

48231

HOSPITAL NAME	INTERMEDIATE REFERENCE LAB	COMPLAINT NUMBER	PB25109
HOSPITAL TYPE	TB Hospital	COMPLAINT RECEIVED DATE/TIME	5/10/2024
DISTRICT	PATIALA	COMPLAINT ATTENDED DATE/TIME	7/10/2024
USER CONTACT DETAILS	7889917139	COMPLAINT RECTIFIED DATE/TIME	14/10/2024
EQUIPMENT TYPE (TICK THE BOX)	<input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER: 7140260050	MANUFACTURER: HAIN LIFE SCIENCE
	EQUIPMENT NAME: G.T. BLOT 48	MODEL NO: G.T. BLOT 48
	DEPARTMENT: NTEP	SERIAL NUMBER: 1095-3

PROBLEM IDENTIFIED: Machine not working.

ACTION TAKEN: As per OEM, this unit has reached its probable life and it is confirmed that the repair of the same is not cost effective as the machine is quite old. It is also informed to you that there is no buy back cost for the same. dated - 14/10/24.

PHASE - NEUTRAL (VOLT):
 PHASE - EARTH (VOLT):
 NEUTRAL - EARTH (VOLT):

ENGINEER'S REMARKS: Suggested the User for the condemnation of this equipment as per policy.

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.
	Motor.	1	/
	Housing & Cables.	1	/

CUSTOMER REMARKS IF ANY:

JOB COMPLETED SATISFACTORY
 YES NO

CUSTOMER SIGNATURE: *Sandeep Raina*
SANDEEP RAINA
 TECHNICAL OFFICER/TECHNICAL MANAGER

FOR AOV INTERNATIONAL LLP
Shubham
SHUBHAM SHARMA
 ENGINEER NAME & SIGNATURE