



AOV INTERNATIONAL LLP
BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF
PUNJAB HEALTH SYSTEMS CORPORATION
SERVICE REPORT



53399

HOSPITAL NAME	CHC Manswala	COMPLAINT NUMBER	P8755210
HOSPITAL TYPE	CHC	COMPLAINT RECEIVED DATE/TIME	18/10/24
DISTRICT	Amritsar	COMPLAINT ATTENDED DATE/TIME	18/10/24
USER CONTACT DETAILS	89468 35577	COMPLAINT RECTIFIED DATE/TIME	18/10/24
EQUIPMENT TYPE (TICK THE BOX) <input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CORRUPTED			

EQUIPMENT DETAIL	EQUIPMENT NUMBER	410008030	MANUFACTURER	oshans
	EQUIPMENT NAME	Cryo table	MODEL NO	NA
	DEPARTMENT	Laboratory	SERIAL NUMBER	NA

PROBLEM IDENTIFIED: machine not working

ACTION TAKEN: Attend the case found that front side table legs is rusted and one back side leg is also rusted. lower body table is also

PHASE - NEUTRAL (VOLT) PHASE - EARTH (VOLT) NEUTRAL - EARTH (VOLT)

ENGINEER'S REMARKS: rusted not be repairable and for safety purpose not be good for patient so this machine recommended to condemnation

SPARES CONSUMED	PART NAME	QTY	SERIAL NO / LOT NO
	/		
/			

CUSTOMER REMARKS IF ANY:

JOB COMPLETED SATISFACTORY YES NO

CUSTOMER SIGNATURE WITH SEAL: *Emergency Medical Officer*
CHC Manswala
Amritsar

FOR AOV INTERNATIONAL LLP

ENGINEER NAME & SIGNATURE: *Navneet*
921144456









