



BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF  
PUNJAB HEALTH SYSTEMS CORPORATION



# INCIDENT REPORT

I. R. No. **2307**  
 Date **26-7-24**  
 Region **Jalandhar**

1 Device Identification	
a. Asset No.	4110040035
b. Asset description	Autoclave (V)
c. Manufacturer Name	REICO
d. Model	MA
e. Serial No.	MA
f. Any other Details	

2 Date & Location	
a. Hospital Name	CHC Parohat
b. Department Name	Labours Room
c. Date of Occurrence	26-7-24
d. Time of Occurrence	10:00AM
e. Date of Reported	26-7-24
f. Time Reported	10:00AM

3 Follow Up Action	
A. BM Work Order No.:	PM29096

3 Damage to Equipment			
A	Major	<input checked="" type="checkbox"/>	B
	Minor	<input type="checkbox"/>	

5 Description of the Incident (State exactly what was the sequence of events leading up to the incident and where the incident occurred Please add another paper if space provided is not enough)

checked And Found that body of Autoclave blast due to over pressure. Now machine is in non-repairable condition. Hence, recommend for condemned.

6 Cause of Incident

a. Physical Damage	<input type="checkbox"/>	j. Inadequate Illumination	<input type="checkbox"/>
b. Missing Spare/Parts	<input type="checkbox"/>	k. Hazardous Method or Procedure	<input type="checkbox"/>
c. Operating without Authorization	<input type="checkbox"/>	l. Non Use of Personal Protective Equipment	<input type="checkbox"/>
d. Wrong Usage / Setting	<input type="checkbox"/>	m. Patient Action	<input type="checkbox"/>
e. Unsafe Equipment	<input checked="" type="checkbox"/>	n. Fire, Explosion or Natural Calamities	<input type="checkbox"/>
f. Unsafe Position or Posture	<input type="checkbox"/>	o. Hazardous Personal Attire	<input type="checkbox"/>
g. Working on Moving or Dangerous Equipment	<input type="checkbox"/>	p. Unsafe Design or Arrangement	<input type="checkbox"/>
h. Distraction, Teasing or Willfull Misconduct	<input type="checkbox"/>	q. Equipment Malfunction or Device Failure	<input type="checkbox"/>
i. Not Guarded or Temporarily Guarded	<input type="checkbox"/>	Other Explain	<input type="checkbox"/>

7 Action Taken on Equipment Repair

a) Minor problem and solved without any financial implications.	<input type="checkbox"/>
b) Minor problem and submitted the quote for required parts.	<input type="checkbox"/>
c) Major problem, needs vendor support for further trouble shooting.	<input type="checkbox"/>
d) Major problem. Can not be repaired, Proposed for BER	<input checked="" type="checkbox"/>

E) Replacement required parts details		Part Nos	Qty.	Approx. Cost
a)				
b)				
c)				
d)				
e)				
Total approx. estimate cost of Repair				

8 Remarks (if any)

1 received Keyon 26/7/24

*(Signature)*  
 Engineer Name **MANRAJ SIMHA**  
 Date **26-7-24**

*(Signature)*  
 NODAL OFFICER  
 Date

*(Signature)*  
 Senior Medical Officer  
 CHC Parohat (KPT)  
 26/7/24

**AOV**



AOV INTERNATIONAL LLP

6215

BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF  
PUNJAB HEALTH SYSTEMS CORPORATION



SERVICE REPORT

HOSPITAL NAME	CHC Pandat	COMPLAINT NUMBER	PB22096
HOSPITAL TYPE	CHC	COMPLAINT RECEIVED DATE/ TIME	24-7-24
DISTRICT	Kapurthala	COMPLAINT ATTENDED DATE/ TIME	26-7-24
USER CONTACT DETAILS	99140 806 43	COMPLAINT RECTIFIED DATE/ TIME	26-7-24
EQUIPMENT TYPE (TICK THE BOX)	<input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CNC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER: 41100110035	MANUFACTURER: Arico
	EQUIPMENT NAME: Autoclave CVJ	MODEL NO: NA
	DEPARTMENT: Labour Room	SERIAL NUMBER: NA

PROBLEM IDENTIFIED: \* Not Working


ACTION TAKEN: \* Checked And Found that Autoclave blast Due to over Pressure. Now the Machine is in Non Repairable condition. Hence Recommend for condemned.

PHASE - NEUTRAL (VOLT): 231V  
 PHASE - EARTH (VOLT): 230V  
 NEUTRAL - EARTH (VOLT): 01V

ENGINEER'S REMARKS: \* Autoclave Recommend for condemned.

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.

CUSTOMER REMARKS IF ANY

JOB COMPLETED SATISFACTORY  YES    NO  
CUSTOMER SIGNATURE WITH SEAL: *[Signature]*  
Senior Medical Officer  
MC P.H. Panchhat (KPT)  
Received  
Heights  
26/7/24  
FOR AOV INTERNATIONAL LLP  
ENGINEER NAME & SIGNATURE: *[Signature]*  
MAHARAJ SIMAN  






SLAVE