



Registration No.  
QAC/R91/0965

# vishal SURGICAL EQUIPMENT CO.

(An ISO 9001 : 2015 Certified Company)

IMPORTERS AND TRADERS OF SURGICAL EQUIPMENTS & INSTRUMENTS

Post Box-2822, B-5, Rajendra Park, Pusa Road, New Delhi - 110060  
Ph. : 25722407, 25732959, 25768165 Telefax No. : 91-11-25752761  
E-mail : vishaldl@vishalsurgical.co.in, vishalsurgicaldl@gmail.com



"Condemnation Report"

Ref. No. ....

Dated 24/10/24

S.M.O,  
Civil Hospital,  
Mohali (PB).


Equipment → Laparoscopic set — 01 No.

Rectification → checked the above said  
equipment properly and found the  
same below fault :-

1. Equipment is used for more than its life  
of ten years.

2. Model was outdated.

Remarks → So treat it as condemnation.

  
Anshu Kumar  
For Vishal Surgical Equipment Co.





**AOV INTERNATIONAL LLP**  
**BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF**  
**PUNJAB HEALTH SYSTEMS CORPORATION**  
**SERVICE REPORT**



44475

HOSPITAL NAME	DH MOHALI	COMPLAINT NUMBER	PB 21347
HOSPITAL TYPE	DH	COMPLAINT RECEIVED DATE/ TIME	11/11/2024
DISTRICT	MOHALI	COMPLAINT ATTENDED DATE/ TIME	11/11/2024
USER CONTACT DETAILS	9872904467	COMPLAINT RECTIFIED DATE/ TIME	11/11/2024

EQUIPMENT TYPE (TICK THE BOX)  UNDER CONTRACT  UNDER WARRANTY  UNDER AMC/CMC  PHYSICAL DAMAGE  CONDEMNED

EQUIPMENT NUMBER :	21600 70 283	MANUFACTURER :	Karl Storz
EQUIPMENT NAME :	Laparoscope set	MODEL NO :	halogen 150
DEPARTMENT :	MCH OT	SERIAL NUMBER :	P221521-B

**PROBLEM IDENTIFIED**  
 Parts not available this machine model was absolute.

**ACTION TAKEN**  
 This machine is not separable, model was absolute. So suggest for Condemned, Equments in used for more than its life of ten year.

PHASE - NEUTRAL (VOLT):  PHASE - EARTH (VOLT):  NEUTRAL - EARTH (VOLT):

**ENGINEER'S REMARKS**  
 This machine is not separable, model was absolute. Parts not available in this machine, Accessories of this machine is faulty. So it is Condemned.

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.

**CUSTOMER REMARKS IF ANY**

IS COMPLETED SATISFACTORY  YES  NO

CUSTOMER SIGNATURE WITH SEAL  
 Dr. Kritika Sharda Shukla  
 SAS Nagar

FOR AOV INTERNATIONAL LLP  
 ENGINEER NAME & SIGNATURE

