



AOV INTERNATIONAL LLP
BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF
PUNJAB HEALTH SYSTEMS CORPORATION



SERVICE REPORT

54968

HOSPITAL NAME	CHC Helway		
HOSPITAL TYPE	CWC		
DISTRICT	Bathinda		
USER CONTACT DETAILS	6280876020		
EQUIPMENT TYPE (TICK THE BOX)	<input checked="" type="checkbox"/> UNDER CONTRACT	<input type="checkbox"/> UNDER WARRANTY	<input type="checkbox"/> UNDER AMC/CMC


COMPLAINT NUMBER	PB 27665
COMPLAINT RECEIVED DATE/ TIME	29-11-24
COMPLAINT ATTENDED DATE/ TIME	29-11-24
COMPLAINT RECTIFIED DATE/ TIME	29-11-24

EQUIPMENT NUMBER:	4030060020
EQUIPMENT NAME:	X-Ray 411c
DEPARTMENT:	Radiology
MANUFACTURER:	Seimens
MODEL NO.:	Multi mobile 25
SERIAL NUMBER:	

PROBLEM IDENTIFIED
 Not working.


ACTION TAKEN
 Dissected and found the Hockins was not getting on. Open the Unit and found Main Board and Diodes Board are Missing. Also Hockins is very old (20 year)

■ PHASE - NEUTRAL (VOLT): ■ PHASE - EARTH (VOLT): ■ NEUTRAL - EARTH (VOLT):

ENGINEER'S REMARKS	Recommended for Condensation		
SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.
			

CUSTOMER REMARKS IF ANY

JOB COMPLETED SATISFACTORY YES NO

CUSTOMER SIGNATURE WITH SEAL
 29/11/24
MEDICAL OFFICER MEHRAJ (BTL.)

FOR AOV INTERNATIONAL LLP
 Rishi Shah SHARAN
ENGINEER NAME & SIGNATURE
