

AOV INTERNATIONAL LLP

BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF  
PUNJAB HEALTH SYSTEMS CORPORATION



SERVICE REPORT

45894

HOSPITAL NAME	DH Jalandhar	COMPLAINT NUMBER	PB97015
HOSPITAL TYPE	DH	COMPLAINT RECEIVED DATE/ TIME	28-11-24
DISTRICT	Jalandhar	COMPLAINT ATTENDED DATE/ TIME	28-11-24
USER CONTACT DETAILS	9988963289	COMPLAINT RECTIFIED DATE/ TIME	7-12-24
EQUIPMENT TYPE (TICK THE BOX)	<input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER: 210110075	MANUFACTURER: Moller Coeller
	EQUIPMENT NAME: Eye operating microscope	MODEL NO: FS1-12
	DEPARTMENT: Eye OT	SERIAL NUMBER: 313

PROBLEM IDENTIFIED: \* Not Working

ACTION TAKEN: \* checked And found that equipment is more than 12 years old And Now out of spares and service support due to parts obsolescence by the OBM. So machine Recommended for condemned.

PHASE - NEUTRAL (VOLT):     
  PHASE - EARTH (VOLT):     
  NEUTRAL - EARTH (VOLT):

ENGINEER'S REMARKS: \* machine Recommended for condemned.

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.
	/	/	/

CUSTOMER REMARKS IF ANY

JOB COMPLETED SATISFACTORY  YES  NO  
 CUSTOMER SIGNATURE WITH SEAL: *[Signature]*  
 FOR AOV INTERNATIONAL LLP: MANRAJ SINGH, 9289640075  
 ENGINEER NAME & SIGNATURE





**BioMedix**  
Optotechnik & Devices Pvt. Ltd.

#258, Amariyothi Layout,  
Domlur, Bangalore - 560071, India.  
Phone: 91-80-49332900  
E-mail: info@biomedixdevices.com  
Web: www.biomedixdevices.com  
CIN: U33112KA1993PTCO14968

**Ref No-**  
**BMOD/HO/SER/Letter/051224/008**  
**Date: 05.12.2024**

**To,**  
**AOV International LLP,**  
**Jalandhar**

**Subject: Condemnation Certificate for Moller Wedel Operating Microscope**

Dear Sir,  
We hereby wish to certify that the **Moller Wedel Operating Microscope S/N 1111**,  
supplied at your hospital is more than 12 years old equipment now and out of Spares & Service  
support due to parts obsolescence.

**Thanks & Regards**



**Authorised Signatory.**

**Branch Offices:**

Shirin Manzil,  
38, Sitaladevi Mahim,  
Mumbai 400016.  
022-24464351, 24474905  
mumbai@biomedixdevices.com

M-1, Aashirwad Complex,  
D-1, Green Park,  
New Delhi 110016  
011-41756375, 41759376  
delhi@biomedixdevices.com

32, 1st Avenue,  
Ashok Nagar,  
Chennai 600083  
044-24891250, 23714716  
chennai@biomedixdevices.com

86,  
Beltola Road,  
Kolkata 700026  
033-24192119, 24863122  
Kolkata@biomedixdevices.com

21, Ground Floor 'Sunrise Mall'  
Mansi Cross Road, Vastrapur,  
Ahmedabad 380015  
079 - 40305101  
ahmedabad@biomedixdevices.com