



# AOV INTERNATIONAL LLP

BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF  
PUNJAB HEALTH SYSTEMS CORPORATION



52658

## SERVICE REPORT

HOSPITAL NAME	CHC SAROYA	COMPLAINT NUMBER	PB-29272
HOSPITAL TYPE	CHC	COMPLAINT RECEIVED DATE/ TIME	22/01/25
DISTRICT	NAWANSHAR	COMPLAINT ATTENDED DATE/ TIME	22/01/25
USER CONTACT DETAILS	7087555135	COMPLAINT RECTIFIED DATE/ TIME	22/01/25
EQUIPMENT TYPE (TICK THE BOX)	<input type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input checked="" type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER: 4180050060	MANUFACTURER: VESTFROST DENMARK
	EQUIPMENT NAME: DEEP freezer-20	MODEL NO: N/A
	DEPARTMENT: P.P. Unit	SERIAL NUMBER: N/A

PROBLEM IDENTIFIED: now MACHINE is not working

ACTION TAKEN: Checked and found the MACHINE Physical conditions very poor and MACHINE is obsolete model So it's non Repairable. Due to non-Availability of Spares in Market. So now MACHINE Recommended for Condemnation

PHASE - NEUTRAL (VOLT):     
 PHASE - EARTH (VOLT):     
 NEUTRAL - EARTH (VOLT):

ENGINEER'S REMARKS: Deep freezer -2 is not Repairable. So Suggested for Condemnation

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.

CUSTOMER REMARKS IF ANY:

JOB COMPLETED SATISFACTORY  YES  NO

CUSTOMER SIGNATURE WITH SEAL: *[Signature]*  
Senior Medical Officer  
CHC Saroya (S.B.S Nagar)

FOR AOV INTERNATIONAL LLP  
Harpreet Singh  
9289690076  
ENGINEER NAME & SIGNATURE: *[Signature]*





