

AOV INTERNATIONAL LLP



BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF
PUNJAB HEALTH SYSTEMS CORPORATION



52666

SERVICE REPORT

HOSPITAL NAME	PNC AVR	COMPLAINT NUMBER	PB29356
HOSPITAL TYPE	PNC	COMPLAINT RECEIVED DATE/ TIME	24-1-25
DISTRICT	Nawanshahr	COMPLAINT ATTENDED DATE/ TIME	27/1/25
USER CONTACT DETAILS	9878198919	COMPLAINT RECTIFIED DATE/ TIME	27/1/25
EQUIPMENT TYPE (TICK THE BOX)	<input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input checked="" type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER :	5180120006	MANUFACTURER :	OTHERS
	EQUIPMENT NAME :	Autoclave Vertical Small	MODEL NO :	Autoclave
	DEPARTMENT :	LABOR ROOM	SERIAL NUMBER :	N/A

PROBLEM IDENTIFIED
Element, Gasket & Element Housing RUSTED

ACTION TAKEN
Checked This machine & found Autoclave Element Housing Body Rusted & Broken due to which Leakage from element. So for Patient safety purpose we suggest this machine for cond.

PHASE - NEUTRAL (VOLT):
 PHASE - EARTH (VOLT):
 NEUTRAL - EARTH (VOLT):

ENGINEER'S REMARKS
Element, Gasket, Body Rusted. So suggested for Condemnation

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.
	N/A		

CUSTOMER REMARKS IF ANY

JOB COMPLETED SATISFACTORY <input checked="" type="checkbox"/> YES <input type="checkbox"/> NO	FOR AOV INTERNATIONAL LLP Harpreet Singh 9289690076 ENGINEER NAME & SIGNATURE
CUSTOMER SIGNATURE WITH SEAL 	



