

33

AOV INTERNATIONAL LLP MAINTENANCE OF BIOMEDICAL EQUIPMENT SERVICE REPORT



HOSPITAL NAME	Sdh Khanna	COMPLAINT NUMBER	HNF.
HOSPITAL TYPE	Sdh	COMPLAINT RECEIVED DATE / TIME	
DISTRICT	Ludhiana	COMPLAINT ATTENDED DATE / TIME	5-8-2022
USER CONTACT DETAILS		COMPLAINT RECTIFIED DATE / TIME	5-8-2022
EQUIPMENT TYPE (TICK THE BOX)	<input type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	BARCODE NUMBER :	3120320150	MANUFACTURER :	Clarity.
	EQUIPMENT NAME :	ECG Machine	MODEL NO :	ECG 100
	DEPARTMENT :	ECG Room	SERIAL NUMBER :	E1-T687-0249

PROBLEM IDENTIFIED: *machine is not working.*

ACTION TAKEN: *Attd: we checked to found machine printer faulty, Battery faulty. Machine is not repairable condition. to Recommend to Condemnation.*

■ PHASE - NEUTRAL (VOLT): *230* ✓
 ■ PHASE - EARTH (VOLT): *234* ✓
 ■ NEUTRAL - EARTH (VOLT): *2V*

ENGINEER'S REMARKS

SPARES CONSUMED OR REQUIRED	PART NAME	QTY	SERIAL NO. / LOT NO.

CUSTOMER REMARK IF ANY

CUSTOMER SIGNATURE WITH SEAL
Bajinder
 M. S. Basu
 ਸਿਨੀਅਰ ਮੈਡੀਕਲ ਆਫਸਰ
 ਆਲ ਹਸਪਤਾਲ, ਖੰਨਾ

FOR AOV INTERNATIONAL LL
 ENGINEER NAME & SIGNATURE
[Signature]