

AOV INTERNATIONAL LLP

6703

BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF
PUNJAB HEALTH SYSTEMS CORPORATION



SERVICE REPORT

HOSPITAL NAME	DH Jalandhar	COMPLAINT NUMBER	HNF
HOSPITAL TYPE	DH	COMPLAINT RECEIVED DATE/ TIME	
DISTRICT	Jalandhar	COMPLAINT ATTENDED DATE/ TIME	20/3/23
USER CONTACT DETAILS		COMPLAINT RECTIFIED DATE/ TIME	20/3/23
EQUIPMENT TYPE (TICK THE BOX)	<input type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CNC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER:	2100110826	MANUFACTURER:	Binn. B Labomed
	EQUIPMENT NAME:	Binocular Telescope	MODEL NO:	VISION 2000
	DEPARTMENT:	TB Dept.	SERIAL NUMBER:	0003

PROBLEM IDENTIFIED	Not Working Supply Card, Regulator, Lamp Missing
--------------------	--

ACTION TAKEN	Checked This machine & Found Regulator, Supply Card, LAMP, Missing & Eye Pcs, Lenses Faulty. Body Rusted or damaged. So not suitable for faulty use so suggested for condemnation.
--------------	--

<input type="checkbox"/> PHASE - NEUTRAL (VOLT):	<input type="checkbox"/> PHASE - EARTH (VOLT):	<input type="checkbox"/> NEUTRAL - EARTH (VOLT):
--	--	--

ENGINEER'S REMARKS	Suggested to Condemnation
--------------------	---------------------------

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.
	1	1	1

CUSTOMER REMARKS IF ANY	
-------------------------	--

JOB COMPLETED SATISFACTORY <input type="checkbox"/> YES <input type="checkbox"/> NO CUSTOMER SIGNATURE WITH SEAL	FOR AOV INTERNATIONAL LLP ENGINEER NAME & SIGNATURE
---	--