

BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF
PUNJAB HEALTH SYSTEMS CORPORATION

SERVICE REPORT

HOSPITAL NAME	DH - Jalandhar	COMPLAINT NUMBER	
HOSPITAL TYPE	DH	COMPLAINT RECEIVED DATE/ TIME	HNE
DISTRICT	Jalandhar	COMPLAINT ATTENDED DATE/ TIME	
USER CONTACT DETAILS		COMPLAINT RECTIFIED DATE/ TIME	20/3/23
EQUIPMENT TYPE (TICK THE BOX)	☐ UNDER CONTRACT ☐ UNDER WARRANTY ☐ UNDER AMC/CNC ☐ PHYSICAL DAMAGE ☐ CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER :	2100110 898	MANUFACTURER :	Lehmann
	EQUIPMENT NAME :	Binocular Microscope	MODEL NO :	VISION 2000
	DEPARTMENT :	TB Dept	SERIAL NUMBER :	0059

PROBLEM IDENTIFIED	Not looking supply card, Regulator, Eye Res. Nic...
--------------------	---

ACTION TAKEN	Checked this machine & found many part faulty & Body damaged supply card, Regulator, Eye Res. Lens are dirty. Not suitable for use suggested for condemnation
--------------	---

☐ PHASE - NEUTRAL (VOLT):	☐ PHASE - EARTH (VOLT):	☐ NEUTRAL - EARTH (VOLT):
--	--	--

ENGINEER'S REMARKS	Machine suggested for Condemnation
--------------------	------------------------------------

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.

CUSTOMER REMARKS IF ANY	
-------------------------	--

JOB COMPLETED SATISFACTORY ☐ YES ☒ NO CUSTOMER SIGNATURE WITH SEAL 	FOR AOV INTERNATIONAL LLP ENGINEER NAME & SIGNATURE
---	--