

AOV INTERNATIONAL LLP

BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF
PUNJAB HEALTH SYSTEMS CORPORATION



SERVICE REPORT

55779

HOSPITAL NAME	DH MKH Patiala	COMPLAINT NUMBER	PB31571
HOSPITAL TYPE	DH	COMPLAINT RECEIVED DATE/ TIME	12-3-25
DISTRICT	Patiala	COMPLAINT ATTENDED DATE/ TIME	21-3-25
USER CONTACT DETAILS	9878390397	COMPLAINT RECTIFIED DATE/ TIME	21-3-25
EQUIPMENT TYPE (TICK THE BOX)	<input checked="" type="checkbox"/> UNDER CONTRACT <input type="checkbox"/> UNDER WARRANTY <input type="checkbox"/> UNDER AMC/CMC <input type="checkbox"/> PHYSICAL DAMAGE <input type="checkbox"/> CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER:	2140110009	MANUFACTURER:	Drum India Biomedical
	EQUIPMENT NAME:	fully Auto analyser	MODEL NO:	XL-640
	DEPARTMENT:	Lab	SERIAL NUMBER:	64038

PROBLEM IDENTIFIED	Machine is not working.
--------------------	-------------------------

ACTION TAKEN	Check & find machine is old model, parts are faulty, Not replaceable. & also Customer is not in used. Machine Condition is not good.
--------------	---

☒ PHASE - NEUTRAL (VOLT):
 ☒ PHASE - EARTH (VOLT):
 ☒ NEUTRAL - EARTH (VOLT):

ENGINEER'S REMARKS	So is recommended for the Condemnation.
--------------------	---

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.
	1	1	1

CUSTOMER REMARKS IF ANY	
-------------------------	--

JOB COMPLETED SATISFACTORY <input checked="" type="checkbox"/> YES <input type="checkbox"/> NO	FOR AOV INTERNATIONAL LLP Dr. ASHIMA JAIN M.O. (Pathologist) PCMS- MKH, PATIALA
CUSTOMER SIGNATURE WITH SEAL 	ENGINEER NAME & SIGNATURE 92896-9080

