



AOV INTERNATIONAL LLP

BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF
PUNJAB HEALTH SYSTEMS CORPORATION



SERVICE REPORT

51259

HOSPITAL NAME	CHC Kot Ise Khan	COMPLAINT NUMBER	PB 31805
HOSPITAL TYPE	CHC	COMPLAINT RECEIVED DATE/ TIME	18/3/25
DISTRICT	Moga	COMPLAINT ATTENDED DATE/ TIME	24/3/25
USER CONTACT DETAILS	7009813709	COMPLAINT RECTIFIED DATE/ TIME	24/3/25
EQUIPMENT TYPE (TICK THE BOX)	☒ UNDER CONTRACT ☐ UNDER WARRANTY ☐ UNDER AMC/CMC ☐ PHYSICAL DAMAGE ☐ CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER :	4150050077	MANUFACTURER :	Haier
	EQUIPMENT NAME :	Deep freezer - 20	MODEL NO :	MA
	DEPARTMENT :	PP Unit	SERIAL NUMBER :	BE0460E0100P237003

PROBLEM IDENTIFIED	Not working
--------------------	-------------

ACTION TAKEN	Check and found that base side freezer is faulty damaged due to rust, base plate is totally damaged and compressor is also faulty and body is not repairable on
--------------	---

■ PHASE - NEUTRAL (VOLT):	230	■ PHASE - EARTH (VOLT):	228	■ NEUTRAL - EARTH (VOLT):	2
---------------------------	-----	-------------------------	-----	---------------------------	---

ENGINEER'S REMARKS	Recommended for Condemnation
-----------------------	------------------------------

SPARES CONSUMED	PART NAME	QTY.	SERIAL NO. / LOT NO.

CUSTOMER REMARKS IF ANY	
-------------------------------	--

JOB COMPLETED SATISFACTORY ☒ YES ☐ NO	MEDICAL OFFICER CHC KOT ISE KHAN MOGA	FOR AOV INTERNATIONAL LLP monpreet Singh 8448985219
CUSTOMER SIGNATURE WITH SEAL	Signature of Medical Officer	ENGINEER NAME & SIGNATURE











