

AOV INTERNATIONAL LLP

BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF
PUNJAB HEALTH SYSTEMS CORPORATION



SERVICE REPORT

55812

HOSPITAL NAME	Knowl TB Hospital	COMPLAINT NUMBER	PB31718
HOSPITAL TYPE	TB Hospital	COMPLAINT RECEIVED DATE/TIME	15/3/25
DISTRICT	Pullahari	COMPLAINT ATTENDED DATE/TIME	9/4/25
USER/CONTACT DETAILS	9988464735	COMPLAINT RECTIFIED DATE/TIME	9/4/25
EQUIPMENT TYPE (TICK THE BOX)	☒ UNDER CONTRACT ☐ UNDER WARRANTY ☐ UNDER AMC/OMC ☐ PHYSICAL DAMAGE ☐ CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER	7140580023	MANUFACTURER	Resmed
	EQUIPMENT NAME	Bipap Machine	MODEL NO	Stellar 100
	DEPARTMENT	Emergency	SERIAL NUMBER	22191269474

PROBLEM IDENTIFIED	Machine is not working.
--------------------	-------------------------

ACTION TAKEN	Checked & found Motor, main PCB, keypad display is not working, gas not repairable condition. So recommended to
--------------	---

PHASE - NEUTRAL (VOLT)	230V	PHASE - EARTH (VOLT)	206V	NEUTRAL - EARTH (VOLT)	4V
------------------------	------	----------------------	------	------------------------	----

ENGINEER'S REMARKS	The Condemnation.
--------------------	-------------------

SPARES CONSUMED	PART NAME	QTY	SERIAL NO / LOT NO.

CUSTOMER REMARKS IF ANY	
-------------------------	--

JOB COMPLETED SATISFACTORY	☒ YES ☐ NO	FOR AOV INTERNATIONAL LLP
----------------------------	---	---------------------------

CUSTOMER SIGNATURE WITH SEAL	 Junior Resident	 92896-4082 ENGINEER NAME & SIGNATURE
------------------------------	---------------------	---

