

AOV INTERNATIONAL LLP

BIO-MEDICAL EQUIPMENT MAINTENANCE & MANAGEMENT PROGRAM (BMMP) OF
PUNJAB HEALTH SYSTEMS CORPORATION



SERVICE REPORT

59638

HOSPITAL NAME	D4 Bathinda	COMPLAINT NUMBER	PB32882
HOSPITAL TYPE	D4	COMPLAINT RECEIVED DATE/ TIME	15-4-25
DISTRICT	Bathinda	COMPLAINT ATTENDED DATE/ TIME	17-4-25
USER CONTACT DETAILS	9803464787	COMPLAINT RECTIFIED DATE/ TIME	17-4-25
EQUIPMENT TYPE (TICK THE BOX)	☐ UNDER CONTRACT ☐ UNDER WARRANTY ☐ UNDER AMC/CMC ☐ PHYSICAL DAMAGE ☐ CONDEMNED		

EQUIPMENT DETAIL	EQUIPMENT NUMBER :	2030100389	MANUFACTURER :	Toul-Vac
	EQUIPMENT NAME :	Suction M/C	MODEL NO :	ft-galaxy
	DEPARTMENT :	OT	SERIAL NUMBER :	N/A

PROBLEM
IDENTIFIED

Not working.

ACTION TAKEN

checked the Machine and found its Suction
Jas, Meter, Pressure Meter, Tubing All are
faulty so we recommended for Condemnation

PHASE - NEUTRAL (VOLT): 221 ■ PHASE - EARTH (VOLT): 221 ■ NEUTRAL - EARTH (VOLT): 2.7

ENGINEER'S
REMARKS

RBER (Recommended for Beyond Economic Repair)

SPARES
CONSUMED

PART NAME

QTY.

SERIAL NO. / LOT NO.

CUSTOMER
REMARKS
IF ANY

JOB COMPLETED SATISFACTORY

MEDICAL OFFICER
CIVIL HOSPITAL
BATHINDA

☒ YES

☐ NO

FOR AOV INTERNATIONAL LLP

Krishan SARKAN

CUSTOMER SIGNATURE WITH SEAL

ENGINEER NAME & SIGNATURE



